

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092418\
 Data File : PL040258.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2018 18:29
 Operator : AJ\SJ
 Sample : J4993-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 NS-OILY-WATER

Manual Integrations
 APPROVED

Sohil
 9/25/2018 3:49:51 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 00:54:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092218.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 22 05:31:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.246	3.841	96931311	30572540	12.741m	9.362m#
2) SA Decachlor...	7.868	8.793	78330142	27215452	8.845m	8.506

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092418\
Data File : PL040258.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2018 18:29
Operator : AJ\SJ
Sample : J4993-06
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
NS-OILY-WATER


Manual Integrations
APPROVED

Sohil
9/25/2018 3:49:51 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 00:54:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092218.M
Quant Title : GC Extractables
QLast Update : Sat Sep 22 05:31:14 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.246 min
Delta R.T.: -0.003 min
Response: 96931311
Conc: 12.74 ng/ml m

Instrument :
ECD_L
ClientSampleId :
NS-OILY-WATER


Manual Integrations
APPROVED

Sohil
9/25/2018 3:49:51 PM




#1 Tetrachloro-m-xylene

R.T.: 3.841 min
Delta R.T.: -0.002 min
Response: 30572540
Conc: 9.36 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.868 min
Delta R.T.: -0.009 min
Response: 78330142
Conc: 8.85 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.793 min
Delta R.T.: -0.003 min
Response: 27215452
Conc: 8.51 ng/ml