

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092420\
 Data File : PL061755.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2020 09:08
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 01:47:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 23 13:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	4.007	85785006	127.4E6	50.285	46.681
2)	SA Decachlor...	8.210	9.094	78081853	110.1E6	47.998	50.029
Target Compounds							
2)	A alpha-BHC	3.894	4.396	125.7E6	191.1E6	51.577	47.834
3)	MA gamma-BHC...	4.177	4.682	113.0E6	176.1E6	51.644	47.565
4)	MA Heptachlor	4.479	5.193	104.9E6	163.3E6	51.397	47.148
5)	MB Aldrin	4.726	5.498	93247204	156.2E6	51.680	48.221
6)	B beta-BHC	4.429	4.853	47155680	71587101	51.631	47.410
7)	B delta-BHC	4.632	5.070	107.0E6	149.7E6	52.206	58.734
8)	B Heptachlo...	5.169	5.886	89747617	143.6E6	51.406	57.956
9)	A Endosulfan I	5.508	6.244	83326283	132.7E6	51.422	48.029
10)	B gamma-Chl...	5.396	6.122	89380959	144.0E6	51.467	49.036
11)	B alpha-Chl...	5.455	6.194	89378982	143.4E6	51.454	47.759
12)	B 4,4'-DDE	5.620	6.352	81201759	130.4E6	51.666	47.780
13)	MA Dieldrin	5.751	6.502	86592736	135.4E6	51.485	47.736
14)	MA Endrin	6.007	6.720	76262274	108.5E6	47.927	44.467
15)	B Endosulfa...	6.280	6.930	79403358	116.5E6	49.111	45.077
16)	A 4,4'-DDD	6.135	6.843	69084533	103.8E6	52.178	46.675
17)	MA 4,4'-DDT	6.376	7.143	68588868	104.3E6	54.372	47.635
18)	B Endrin al...	6.448	7.054	69442016	102.1E6	49.383	47.680
19)	B Endosulfa...	6.661	7.278	76374642	111.2E6	49.368	47.474
20)	A Methoxychlor	6.921	7.603	41337964	60479551	51.128	48.001
21)	B Endrin ke...	7.151	7.750	90204293	118.5E6	52.782	50.495
22)	Mirex	7.340	8.205	68069870	87315407	48.555	49.556



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092420\
Data File : PL061755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2020 09:08
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 01:47:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
Quant Title : GC Extractables
QLast Update : Wed Sep 23 13:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.458 min
Delta R.T.: 0.000 min
Response: 85785006
Conc: 50.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 127356860
Conc: 46.68 ng/ml




#2 alpha-BHC

R.T.: 3.894 min
Delta R.T.: 0.000 min
Response: 125714468
Conc: 51.58 ng/ml



#2 alpha-BHC


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 191065439
Conc: 47.83 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.177 min
Delta R.T.: 0.000 min
Response: 112957731
Conc: 51.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.682 min
Delta R.T.: 0.000 min
Response: 176082828
Conc: 47.57 ng/ml




#4 Heptachlor

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 104883304
Conc: 51.40 ng/ml



#4 Heptachlor


R.T.: 5.193 min
Delta R.T.: 0.000 min
Response: 163267878
Conc: 47.15 ng/ml



#5 Aldrin


R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 93247204
Conc: 51.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.498 min
Delta R.T.: 0.000 min
Response: 156209756
Conc: 48.22 ng/ml




#6 beta-BHC

R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 47155680
Conc: 51.63 ng/ml



#6 beta-BHC


R.T.: 4.853 min
Delta R.T.: 0.000 min
Response: 71587101
Conc: 47.41 ng/ml



#7 delta-BHC


R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 107034506
Conc: 52.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 5.070 min
Delta R.T.: 0.000 min
Response: 149660150
Conc: 58.73 ng/ml




#8 Heptachlor epoxide

R.T.: 5.169 min
Delta R.T.: 0.000 min
Response: 89747617
Conc: 51.41 ng/ml



#8 Heptachlor epoxide


R.T.: 5.886 min
Delta R.T.: 0.000 min
Response: 143636692
Conc: 57.96 ng/ml



#9 Endosulfan I


R.T.: 5.508 min
Delta R.T.: 0.000 min
Response: 83326283
Conc: 51.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.244 min
Delta R.T.: 0.000 min
Response: 132681284
Conc: 48.03 ng/ml




#10 gamma-Chlordane

R.T.: 5.396 min
Delta R.T.: 0.000 min
Response: 89380959
Conc: 51.47 ng/ml



#10 gamma-Chlordane


R.T.: 6.122 min
Delta R.T.: 0.000 min
Response: 144002131
Conc: 49.04 ng/ml



#11 alpha-Chlordane


R.T.: 5.455 min
Delta R.T.: 0.000 min
Response: 89378982
Conc: 51.45 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.194 min
Delta R.T.: 0.000 min
Response: 143446451
Conc: 47.76 ng/ml




#12 4,4'-DDE

R.T.: 5.620 min
Delta R.T.: 0.000 min
Response: 81201759
Conc: 51.67 ng/ml



#12 4,4'-DDE


R.T.: 6.352 min
Delta R.T.: 0.000 min
Response: 130430405
Conc: 47.78 ng/ml



#13 Dieldrin


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 86592736
Conc: 51.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.502 min
Delta R.T.: 0.000 min
Response: 135424568
Conc: 47.74 ng/ml




#14 Endrin

R.T.: 6.007 min
Delta R.T.: 0.000 min
Response: 76262274
Conc: 47.93 ng/ml



#14 Endrin


R.T.: 6.720 min
Delta R.T.: 0.000 min
Response: 108482334
Conc: 44.47 ng/ml



#15 Endosulfan II


R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 79403358
Conc: 49.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.930 min
Delta R.T.: -0.001 min
Response: 116509573
Conc: 45.08 ng/ml




#16 4,4'-DDD

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 69084533
Conc: 52.18 ng/ml



#16 4,4'-DDD


R.T.: 6.843 min
Delta R.T.: 0.000 min
Response: 103763044
Conc: 46.68 ng/ml



#17 4,4'-DDT


R.T.: 6.376 min
Delta R.T.: 0.000 min
Response: 68588868
Conc: 54.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 104311062
Conc: 47.63 ng/ml




#18 Endrin aldehyde

R.T.: 6.448 min
Delta R.T.: 0.000 min
Response: 69442016
Conc: 49.38 ng/ml



#18 Endrin aldehyde


R.T.: 7.054 min
Delta R.T.: 0.000 min
Response: 102057413
Conc: 47.68 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.661 min
Delta R.T.: 0.000 min
Response: 76374642
Conc: 49.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.278 min
Delta R.T.: 0.000 min
Response: 111166198
Conc: 47.47 ng/ml




#20 Methoxychlor

R.T.: 6.921 min
Delta R.T.: -0.001 min
Response: 41337964
Conc: 51.13 ng/ml



#20 Methoxychlor


R.T.: 7.603 min
Delta R.T.: 0.000 min
Response: 60479551
Conc: 48.00 ng/ml



#21 Endrin ketone


R.T.: 7.151 min
Delta R.T.: 0.000 min
Response: 90204293
Conc: 52.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.750 min
Delta R.T.: 0.000 min
Response: 118513639
Conc: 50.50 ng/ml




#22 Mirex

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 68069870
Conc: 48.56 ng/ml



#22 Mirex


R.T.: 8.205 min
Delta R.T.: -0.001 min
Response: 87315407
Conc: 49.56 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.210 min
Delta R.T.: -0.002 min
Response: 78081853
Conc: 48.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 9.094 min
Delta R.T.: 0.000 min
Response: 110131668
Conc: 50.03 ng/ml