

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092622\
 Data File : PL077914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2022 12:17
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 00:13:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 13 13:26:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.415	2.624	78522151	133.0E6	48.985	44.785
28)	SA Decachlor...	8.909	7.724	61733106	102.6E6	47.765	40.886
Target Compounds							
2)	A alpha-BHC	3.880	3.116	118.0E6	206.7E6	49.706	45.587
3)	MA gamma-BHC...	4.216	3.441	109.8E6	186.4E6	49.632	46.170
4)	MA Heptachlor	4.792	3.772	111.8E6	179.6E6	50.496	45.339
5)	MB Aldrin	5.133	4.046	100.2E6	172.8E6	50.399	45.538
6)	B beta-BHC	4.429	3.741	48346294	75001650	49.745	45.051
7)	B delta-BHC	4.673	3.963	102.9E6	184.7E6	50.139	45.524
8)	B Heptachlo...	5.566	4.545	90585673	153.1E6	51.962	45.287
9)	A Endosulfan I	5.953	4.910	81471157	142.4E6	50.550	45.128
10)	B gamma-Chl...	5.822	4.794	88638782	154.0E6	50.625	45.184
11)	B alpha-Chl...	5.899	4.857	87544882	149.6E6	50.682	45.177
12)	B 4,4'-DDE	6.078	5.051	81630081	147.9E6	51.280	45.447
13)	MA Dieldrin	6.230	5.174	81138803	150.9E6	50.271	45.503
14)	MA Endrin	6.462	5.447	69252297	131.3E6	50.074	46.060
15)	B Endosulfa...	6.691	5.743	66698494	120.1E6	48.284	45.576
16)	A 4,4'-DDD	6.604	5.604	60994300	119.8E6	52.094	46.234
17)	MA 4,4'-DDT	6.914	5.853	64853118	122.1E6	51.181	46.354
18)	B Endrin al...	6.821	5.922	53046436	92792999	50.163	44.968
19)	B Endosulfa...	7.055	6.146	63822321	120.2E6	49.838	44.703
20)	A Methoxychlor	7.396	6.430	33420769	62204917	50.603	45.424
21)	B Endrin ke...	7.544	6.645	67568054	134.5E6	48.899	44.423
22)	Mirex	8.004	6.821	51394910	95779311	48.948	44.551

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092622\
Data File : PL077914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2022 12:17
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 00:13:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
Quant Title : GC Extractables
QLast Update : Tue Sep 13 13:26:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

