

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092718\
 Data File : PL040423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2018 12:10
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 9/28/2018 5:42:56 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 02:30:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 27 01:27:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.211	3.807	389.0E6	158.8E6	53.139	50.296
28)	SA Decachlor...	7.814	8.731	500.2E6	152.9E6	52.617	48.297
Target Compounds							
2)	A alpha-BHC	3.633	4.178	631.0E6	239.8E6	53.933	50.585
3)	MA gamma-BHC...	3.903	4.452	590.5E6	223.2E6	54.307	49.178
4)	MA Heptachlor	4.177	4.951	632.2E6	220.1E6	53.726	51.812
5)	MB Aldrin	4.409	5.247	580.1E6	208.8E6	54.901	49.485
6)	B beta-BHC	4.157	4.615	215.2E6	96557915	50.986	48.324
7)	B delta-BHC	4.343	4.822	601.9E6	222.3E6	53.925	48.158m
8)	B Heptachlor...	4.840	5.623	551.4E6	192.7E6	53.273	50.207
9)	A Endosulfan I	5.164	5.971	540.5E6	180.8E6	53.622	48.593
10)	B gamma-Chl...	5.058	5.853	588.1E6	197.5E6	54.152	49.985
11)	B alpha-Chl...	5.113	5.925	565.0E6	192.8E6	53.932	49.120
12)	B 4,4'-DDE	5.283	6.084	593.3E6	194.7E6	53.679	48.363
13)	MA Dieldrin	5.399	6.221	584.3E6	188.9E6	53.344	49.453
14)	MA Endrin	5.648	6.433	485.8E6	147.0E6	48.900	45.005
15)	B Endosulfa...	5.922	6.637	534.7E6	165.4E6	53.674	47.311
16)	A 4,4'-DDD	5.788	6.562	473.4E6	149.9E6	51.280	47.197
17)	MA 4,4'-DDT	6.017	6.858	430.3E6	150.3E6	53.205m	49.337
18)	B Endrin al...	6.089	6.759	443.4E6	146.7E6	52.805	49.016
19)	B Endosulfa...	6.298	6.980	498.2E6	154.3E6	52.055	47.964
20)	A Methoxychlor	6.565	7.314	247.1E6	75726029	51.165	49.691
21)	B Endrin ke...	6.774	7.438	594.8E6	160.0E6	54.824	52.057

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092718\
Data File : PL040423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2018 12:10
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
9/28/2018 5:42:56 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 02:30:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
Quant Title : GC Extractables
QLast Update : Thu Sep 27 01:27:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

