

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092724\
 Data File : PL092134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2024 20:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/29/2024
 Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 00:27:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092324.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 23 16:44:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.781	109.3E6	138.1E6	49.141	55.073
28)	SA Decachlor...	9.060	7.922	74873049	117.8E6	45.704	47.076
Target Compounds							
2)	A alpha-BHC	3.998	3.284	154.7E6	204.7E6	49.947	56.173
3)	MA gamma-BHC...	4.331	3.614	155.3E6	194.3E6	50.102	55.026
4)	MA Heptachlor	4.920	3.954	129.2E6	179.1E6	46.725	51.907
5)	MB Aldrin	5.261	4.232	131.2E6	182.8E6	47.220	53.746m
6)	B beta-BHC	4.528	3.914	62746451	80041239	47.307	53.251
7)	B delta-BHC	4.775	4.143	141.1E6	199.0E6	48.906	57.317
8)	B Heptachlo...	5.687	4.736	121.3E6	158.4E6	47.661m	52.404
9)	A Endosulfan I	6.072	5.106	102.6E6	124.3E6	44.333m	45.292
10)	B gamma-Chl...	5.942	4.987	115.6E6	159.4E6	46.768m	52.990
11)	B alpha-Chl...	6.021	5.050	116.5E6	155.2E6	47.299m	51.838
12)	B 4,4'-DDE	6.196	5.238	101.7E6	150.0E6	46.991	53.978m
13)	MA Dieldrin	6.349	5.371	112.0E6	156.5E6	45.912	53.372
14)	MA Endrin	6.578	5.647	88247412	142.2E6	42.824	54.234 #
15)	B Endosulfa...	6.798	5.941	104.1E6	129.8E6	47.318	51.881
16)	A 4,4'-DDD	6.712	5.794	85469091	122.1E6	47.596m	57.149
17)	MA 4,4'-DDT	7.027	6.044	80771244	113.3E6	43.869	48.971
18)	B Endrin al...	6.927	6.120	81793792	104.1E6	46.933	51.000
19)	B Endosulfa...	7.162	6.343	88764256	122.0E6	44.779	50.559
20)	A Methoxychlor	7.503	6.620	40920075	60214118	41.682	48.158
21)	B Endrin ke...	7.647	6.847	97382029	150.7E6	44.314	54.411m
22)	Mirex	8.121	7.028	78636200	114.9E6	43.663	47.231m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092724\
Data File : PL092134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2024 20:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 09/29/2024

Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 00:27:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092324.M
Quant Title : GC Extractables
QLast Update : Mon Sep 23 16:44:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

