

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092818\
 Data File : PL040508.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2018 16:32
 Operator : AJ\SJ
 Sample : PB113334TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 22:46:56 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 27 01:27:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.210	3.806	171.8E6	63475314	23.475	20.101
28)	SA Decachlor...	7.813	8.730	232.7E6	69626470	24.481	21.986
Target Compounds							
3)	MA gamma-BHC...	3.952f	0.000	55452947	0	5.100	N.D. #
7)	B delta-BHC	0.000	4.845	0	5436104	N.D.	1.178 #
8)	B Heptachlo...	4.892f	0.000	38298252	0	3.700	N.D. #
15)	B Endosulfa...	5.867f	0.000	26934422	0	2.704	N.D. #
16)	A 4,4'-DDD	0.000	6.515f	0	24005611	N.D.	7.556 #
20)	A Methoxychlor	6.496f	0.000	66168737	0	13.699	N.D. #
27)	Chlordane-5	5.867	0.000	26934422	0	66.726	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092818\
Data File : PL040508.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2018 16:32
Operator : AJ\SJ
Sample : PB113334TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 22:46:56 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
Quant Title : GC Extractables
QLast Update : Thu Sep 27 01:27:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.210 min
Delta R.T.: -0.001 min
Response: 171834753
Conc: 23.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.001 min
Response: 63475314
Conc: 20.10 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 3.952 min
Delta R.T.: 0.049 min
Response: 55452947
Conc: 5.10 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T.: 4.452 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.342 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.845 min
Delta R.T.: 0.021 min
Response: 5436104
Conc: 1.18 ng/ml




#8 Heptachlor epoxide

R.T.: 4.892 min
Delta R.T.: 0.053 min
Response: 38298252
Conc: 3.70 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.623 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 5.867 min
Delta R.T.: -0.055 min
Response: 26934422
Conc: 2.70 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.637 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 5.787 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.515 min
Delta R.T.: -0.047 min
Response: 24005611
Conc: 7.56 ng/ml



#20 Methoxychlor


R.T.: 6.496 min
Delta R.T.: -0.069 min
Response: 66168737
Conc: 13.70 ng/ml

Instrument :
ECD_L
ClientSampleId :




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 7.314 min
Response: 0
Conc: N.D.




#27 Chlordane-5

R.T.: 5.867 min
Delta R.T.: -0.058 min
Response: 26934422
Conc: 66.73 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T. : 6.710 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.813 min
Delta R.T.: -0.002 min
Response: 232713301
Conc: 24.48 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.730 min
Delta R.T.: -0.001 min
Response: 69626470
Conc: 21.99 ng/ml