

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092819\
 Data File : PL052804.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2019 20:09
 Operator : SG\AJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 05:36:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092819CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 28 05:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	4.048	371.6E6	99433716	39.722	39.248
27)	SA Decachlor...	8.346	9.125	691.7E6	205.2E6	75.845	75.514
Target Compounds							
2)	A alpha-BHC	3.994	4.438	588.8E6	138.3E6	41.006	41.181
3)	MA gamma-BHC...	4.281	4.724	541.4E6	131.4E6	41.208	40.822
4)	MA Heptachlor	4.590	5.234	557.6E6	139.2E6	40.092	39.787
9)	A Endosulfan I	5.632	6.281	448.2E6	112.9E6	40.016	39.012
13)	MA Dieldrin	5.877	6.538	1014.0E6	247.2E6	80.282	80.333
14)	MA Endrin	6.136	6.756	826.3E6	194.6E6	80.040	79.020
16)	A 4,4'-DDD	6.259	6.876	687.3E6	179.6E6	81.299	80.091
17)	MA 4,4'-DDT	6.503	7.175	788.2E6	191.7E6	80.762	79.574
20)	A Methoxychlor	7.049	7.632	1819.0E6	503.1E6	378.129	382.324


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092819\
Data File : PL052804.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2019 20:09
Operator : SG\AJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
INDA401

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 05:36:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092819CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 28 05:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.549 min
Delta R.T.: 0.000 min
Response: 371577506
Conc: 39.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401




#1 Tetrachloro-m-xylene

R.T.: 4.048 min
Delta R.T.: 0.000 min
Response: 99433716
Conc: 39.25 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 588784243
Conc: 41.01 ng/ml



#2 alpha-BHC


R.T.: 4.438 min
Delta R.T.: 0.000 min
Response: 138318995
Conc: 41.18 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.281 min
Delta R.T.: 0.000 min
Response: 541355277
Conc: 41.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401




#3 gamma-BHC (Lindane)

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 131394403
Conc: 40.82 ng/ml




#4 Heptachlor

R.T.: 4.590 min
Delta R.T.: 0.000 min
Response: 557636194
Conc: 40.09 ng/ml



#4 Heptachlor


R.T.: 5.234 min
Delta R.T.: 0.000 min
Response: 139174968
Conc: 39.79 ng/ml



#9 Endosulfan I


R.T.: 5.632 min
Delta R.T.: 0.000 min
Response: 448160084
Conc: 40.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401




#9 Endosulfan I

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 112945098
Conc: 39.01 ng/ml




#13 Dieldrin

R.T.: 5.877 min
Delta R.T.: 0.000 min
Response: 1014030523
Conc: 80.28 ng/ml



#13 Dieldrin


R.T.: 6.538 min
Delta R.T.: 0.000 min
Response: 247220251
Conc: 80.33 ng/ml



#14 Endrin


R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 826326528
Conc: 80.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401




#14 Endrin

R.T.: 6.756 min
Delta R.T.: 0.000 min
Response: 194598737
Conc: 79.02 ng/ml




#16 4,4'-DDD

R.T.: 6.259 min
Delta R.T.: 0.000 min
Response: 687333196
Conc: 81.30 ng/ml



#16 4,4'-DDD


R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 179599160
Conc: 80.09 ng/ml



#17 4,4'-DDT


R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 788192637
Conc: 80.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401




#17 4,4'-DDT

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 191714439
Conc: 79.57 ng/ml




#20 Methoxychlor

R.T.: 7.049 min
Delta R.T.: 0.000 min
Response: 1819000809
Conc: 378.13 ng/ml



#20 Methoxychlor


R.T.: 7.632 min
Delta R.T.: 0.000 min
Response: 503111091
Conc: 382.32 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.346 min
Delta R.T.: 0.000 min
Response: 691674355
Conc: 75.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA401



#27 Decachlorobiphenyl

R.T.: 9.125 min
Delta R.T.: 0.000 min
Response: 205207260
Conc: 75.51 ng/ml