

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX092919.M
 Title : GC Extractables
 Last Update : Sat Sep 28 05:43:38 2019
 Response Via : Initial Calibration

Calibration Files

500 =PL052826.D 1000 =PL052824.D 250 =PL052827.D
 750 =PL052825.D 100 =PL052828.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.078	1.062	1.039	1.050	1.148	1.075	E7	3.99
2)	Toxaphene-1	7.453	7.694	7.273	7.415	8.613	7.690	E4	7.00
3)	Toxaphene-2	1.923	1.948	1.772	1.893	1.934	1.894	E5	3.76
4)	Toxaphene-3	6.517	6.640	5.960	6.525	6.815	6.491	E4	4.94
5)	Toxaphene-4	7.494	7.590	6.863	7.266	7.486	7.340	E4	3.98
6)	Toxaphene-5	1.961	1.964	1.852	1.944	2.052	1.955	E5	3.64
7) SA	Decachlorobiphenyl	1.033	0.991	1.034	0.993	1.233	1.057	E7	9.53

Signal #2 Calibration Files

500 =PL052826.D 1000 =PL052824.D 250 =PL052827.D
 750 =PL052825.D 100 =PL052828.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.826	2.846	2.763	2.780	3.080	2.859	E6	4.47
2)	Toxaphene-1	1.468	1.507	1.554	1.445	1.723	1.540	E4	7.19
3)	Toxaphene-2	1.748	1.777	1.620	1.787	1.159	1.618	E4	16.37
4)	Toxaphene-3	7.077	7.106	7.310	6.875	7.960	7.266	E4	5.75
5)	Toxaphene-4	4.372	4.352	4.272	4.190	4.713	4.380	E4	4.56
6)	Toxaphene-5	2.091	2.078	2.057	1.940	2.225	2.078	E4	4.89
7) SA	Decachlorobiphenyl	3.102	2.992	3.123	2.965	3.657	3.168	E6	8.89

(#) = Out of Range