

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092920\
 Data File : PL061847.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2020 10:37
 Operator : DD\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 01:08:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 23 13:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	4.006	37587253	55516726	22.033	20.349
28)	SA Decachlor...	8.211	9.094	35494352	49012644	21.819	22.265
Target Compounds							
2)	A alpha-BHC	3.894	4.396	24199845	35682225	9.928	8.933
3)	MA gamma-BHC...	4.178	4.681	21368406	32892997	9.770	8.885
6)	B beta-BHC	4.429	4.852	10394418	14438382	11.381	9.562
14)	MA Endrin	6.009	6.720	80074808	115.9E6	50.323	47.523
16)	A 4,4'-DDD	6.136	6.843	2724089	4802948	2.057	2.160
17)	MA 4,4'-DDT	6.377	7.143	143.5E6	215.5E6	113.764	98.431
20)	A Methoxychlor	6.922	7.603	198.4E6	293.1E6	245.414	232.635
21)	B Endrin ke...	7.151	7.750	4222909	3648551	2.471	1.555 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092920\
Data File : PL061847.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2020 10:37
Operator : DD\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 01:08:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
Quant Title : GC Extractables
QLast Update : Wed Sep 23 13:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.458 min
Delta R.T.: 0.000 min
Response: 37587253
Conc: 22.03 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.006 min
Delta R.T.: 0.000 min
Response: 55516726
Conc: 20.35 ng/ml




#2 alpha-BHC

R.T.: 3.894 min
Delta R.T.: 0.000 min
Response: 24199845
Conc: 9.93 ng/ml



#2 alpha-BHC


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 35682225
Conc: 8.93 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.178 min
Delta R.T.: 0.000 min
Response: 21368406
Conc: 9.77 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.681 min
Delta R.T.: 0.000 min
Response: 32892997
Conc: 8.89 ng/ml




#6 beta-BHC

R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 10394418
Conc: 11.38 ng/ml



#6 beta-BHC


R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 14438382
Conc: 9.56 ng/ml



#14 Endrin


R.T.: 6.009 min
Delta R.T.: 0.000 min
Response: 80074808
Conc: 50.32 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 6.720 min
Delta R.T.: -0.002 min
Response: 115937200
Conc: 47.52 ng/ml




#16 4,4'-DDD

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 2724089
Conc: 2.06 ng/ml



#16 4,4'-DDD


R.T.: 6.843 min
Delta R.T.: -0.001 min
Response: 4802948
Conc: 2.16 ng/ml



#17 4,4'-DDT


R.T.: 6.377 min
Delta R.T.: 0.000 min
Response: 143511384
Conc: 113.76 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 215545546
Conc: 98.43 ng/ml




#20 Methoxychlor

R.T.: 6.922 min
Delta R.T.: 0.000 min
Response: 198422324
Conc: 245.41 ng/ml



#20 Methoxychlor


R.T.: 7.603 min
Delta R.T.: 0.000 min
Response: 293114416
Conc: 232.63 ng/ml



#21 Endrin ketone


R.T.: 7.151 min
Delta R.T.: 0.000 min
Response: 4222909
Conc: 2.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 7.750 min
Delta R.T.: 0.000 min
Response: 3648551
Conc: 1.55 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.211 min
Delta R.T.: 0.000 min
Response: 35494352
Conc: 21.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.094 min
Delta R.T.: 0.000 min
Response: 49012644
Conc: 22.26 ng/ml