

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092922\
 Data File : PL077995.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2022 18:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 23:42:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 13 13:26:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.414	2.624	77815715	135.6E6	48.544	45.690
2)	SA Decachlor...	8.902	7.719	62692645	103.2E6	48.508	41.124
Target Compounds							
2)	A alpha-BHC	3.879	3.116	116.3E6	211.5E6	49.008	46.639
3)	MA gamma-BHC...	4.215	3.439	108.0E6	191.1E6	48.818	47.332
4)	MA Heptachlor	4.790	3.771	107.6E6	182.6E6	48.600	46.097
5)	MB Aldrin	5.130	4.044	98831804	178.1E6	49.728	46.944
6)	B beta-BHC	4.428	3.739	47402358	77334478	48.774	46.452
7)	B delta-BHC	4.671	3.962	96984962	189.4E6	47.236	46.687
8)	B Heptachlo...	5.563	4.543	91973522	155.5E6	52.758	45.998
9)	A Endosulfan I	5.950	4.908	82305380	145.1E6	51.067	45.979
10)	B gamma-Chl...	5.819	4.792	88511250	158.0E6	50.552	46.350
11)	B alpha-Chl...	5.896	4.855	87013638	152.9E6	50.375	46.173
12)	B 4,4'-DDE	6.075	5.049	82857095	153.3E6	52.051	47.123
13)	MA Dieldrin	6.227	5.172	81481562	155.7E6	50.483	46.964
14)	MA Endrin	6.459	5.444	68509505	131.2E6	49.537	46.002
15)	B Endosulfa...	6.687	5.740	66443609	122.6E6	48.100	46.520
16)	A 4,4'-DDD	6.600	5.601	63560197	126.0E6	54.285	48.621
17)	MA 4,4'-DDT	6.911	5.850	60767528	117.1E6	47.957	44.451
18)	B Endrin al...	6.818	5.919	54837564	98332061	51.857	47.652
19)	B Endosulfa...	7.051	6.142	63240009	122.7E6	49.383	45.611
20)	A Methoxychlor	7.392	6.427	31167176	59719726	47.191	43.609
21)	B Endrin ke...	7.539	6.641	67524401	137.5E6	48.867	45.404
22)	Mirex	7.999	6.817	50734907	96399465	48.320	44.840

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092922\
Data File : PL077995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2022 18:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 23:42:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
Quant Title : GC Extractables
QLast Update : Tue Sep 13 13:26:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

