

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL093019\
 Data File : PL052835.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2019 09:33
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/2/2019 1:45:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 14:15:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.556	4.047	498.0E6	126.2E6	50.068	46.984
28)	SA Decachlor...	8.351	9.127	499.4E6	138.2E6	49.856	47.117
Target Compounds							
2)	A alpha-BHC	4.000	4.438	819.8E6	174.7E6	51.396	45.501
3)	MA gamma-BHC...	4.288	4.724	743.9E6	163.0E6	51.074	45.015
4)	MA Heptachlor	4.597	5.234	745.5E6	178.2E6	49.890	46.768
5)	MB Aldrin	4.849	5.540	700.6E6	163.3E6	50.936	47.112
6)	B beta-BHC	4.538	4.891	298.2E6	80559013	48.035	51.252m
7)	B delta-BHC	4.747	5.110	725.6E6	156.1E6	50.638	43.127
8)	B Heptachlo...	5.296	5.926	647.4E6	156.1E6	49.678	46.568
9)	A Endosulfan I	5.638	6.282	605.8E6	143.8E6	49.919	45.992
10)	B gamma-Chl...	5.524	6.159	676.9E6	158.7E6	50.331	46.909
11)	B alpha-Chl...	5.584	6.232	659.1E6	157.0E6	49.972	46.630
12)	B 4,4'-DDE	5.747	6.388	590.0E6	151.8E6	50.838	46.466
13)	MA Dieldrin	5.883	6.540	661.7E6	154.4E6	51.259	47.323
14)	MA Endrin	6.142	6.758	572.8E6	130.3E6	50.284	46.926
15)	B Endosulfa...	6.415	6.965	569.1E6	136.4E6	49.665	45.965
16)	A 4,4'-DDD	6.265	6.877	448.7E6	114.7E6	49.918	46.786
17)	MA 4,4'-DDT	6.508	7.176	520.6E6	121.6E6	50.974	46.725
18)	B Endrin al...	6.583	7.089	467.6E6	113.6E6	49.793	46.322
19)	B Endosulfa...	6.797	7.312	479.0E6	125.3E6	46.013	45.893
20)	A Methoxychlor	7.054	7.633	245.6E6	66572811	47.623	47.284
21)	B Endrin ke...	7.291	7.783	568.0E6	129.5E6	48.734	45.378
22)	Mirex	7.486	8.239	428.1E6	104.6E6	45.337	46.526

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL093019\
Data File : PL052835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2019 09:33
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/2/2019 1:45:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 14:15:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

