

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL093019\
 Data File : PL052882.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2019 22:35
 Operator : SG\AJ
 Sample : PB123476BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB123476BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 01:03:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	4.047	215.7E6	60415690	21.684	22.499
28)	SA Decachlor...	8.344	9.122	237.3E6	69949697	23.688	23.843
Target Compounds							
2)	A alpha-BHC	3.993	4.438	868.4E6	209.0E6	54.446	54.416
3)	MA gamma-BHC...	4.281	4.723	799.8E6	197.6E6	54.914	54.562
4)	MA Heptachlor	4.589	5.233	828.8E6	208.1E6	55.466	54.611
5)	MB Aldrin	4.842	5.539	793.3E6	196.2E6	57.669	56.612
6)	B beta-BHC	4.530	4.892	329.4E6	89409653	53.057	56.883
7)	B delta-BHC	4.739	5.109	735.9E6	184.1E6	51.358	50.846
8)	B Heptachlo...	5.288	5.924	720.1E6	191.6E6	55.261	57.175
9)	A Endosulfan I	5.631	6.281	674.0E6	174.6E6	55.536	55.852
10)	B gamma-Chl...	5.517	6.158	734.9E6	195.3E6	54.644	57.717
11)	B alpha-Chl...	5.576	6.231	713.9E6	185.5E6	54.124	55.100
12)	B 4,4'-DDE	5.739	6.386	642.7E6	180.8E6	55.375	55.350
13)	MA Dieldrin	5.876	6.538	727.5E6	179.5E6	56.356	55.025
14)	MA Endrin	6.134	6.756	624.8E6	148.4E6	54.850	53.435
15)	B Endosulfa...	6.407	6.963	600.2E6	160.0E6	52.372	53.940
16)	A 4,4'-DDD	6.257	6.875	506.8E6	137.9E6	56.381	56.241
17)	MA 4,4'-DDT	6.500	7.174	566.6E6	143.3E6	55.486	55.050
18)	B Endrin al...	6.575	7.087	498.9E6	131.8E6	53.123	53.734
19)	B Endosulfa...	6.790	7.310	570.9E6	149.6E6	54.847	54.806
20)	A Methoxychlor	7.047	7.631	276.2E6	82558503	53.541	58.638
21)	B Endrin ke...	7.283	7.781	634.5E6	160.0E6	54.437	56.074
22)	Mirex	7.479	8.236	477.0E6	122.2E6	50.514	54.319

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL093019\
Data File : PL052882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2019 22:35
Operator : SG\AJ
Sample : PB123476BS
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB123476BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:03:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

