

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100118\
 Data File : PL040535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2018 16:58
 Operator : AJ\SJ
 Sample : J4777-37
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 00:59:19 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 27 01:27:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.209	3.806	102.1E6	37249349	13.953	11.796
28)	SA Decachlor...	7.811	8.730	105.1E6	36709737	11.052	11.592
Target Compounds							
7)	B delta-BHC	0.000	4.846	0	4160871	N.D.	0.901 #
12)	B 4,4'-DDE	5.279	6.083	40424733	15470560	3.657	3.842



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100118\
Data File : PL040535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2018 16:58
Operator : AJ\SJ
Sample : J4777-37
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 00:59:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
Quant Title : GC Extractables
QLast Update : Thu Sep 27 01:27:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.209 min
Delta R.T.: -0.002 min
Response: 102138560
Conc: 13.95 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.002 min
Response: 37249349
Conc: 11.80 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.342 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.846 min
Delta R.T.: 0.022 min
Response: 4160871
Conc: 0.90 ng/ml



#12 4,4'-DDE


R.T.: 5.279 min
Delta R.T.: -0.004 min
Response: 40424733
Conc: 3.66 ng/ml

Instrument :
ECD_L
ClientSampleId :




#12 4,4'-DDE

R.T.: 6.083 min
Delta R.T.: -0.002 min
Response: 15470560
Conc: 3.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.811 min
Delta R.T.: -0.004 min
Response: 105054161
Conc: 11.05 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.730 min
Delta R.T.: -0.002 min
Response: 36709737
Conc: 11.59 ng/ml