

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100118\
 Data File : PL040546.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2018 23:20
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 10/2/2018 8:54:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 08:38:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 27 01:27:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.209	3.806	402.8E6	148.6E6	55.034	47.052
28)	SA Decachlor...	7.810	8.729	451.4E6	133.8E6	47.489	42.250
Target Compounds							
2)	A alpha-BHC	3.631	4.178	652.9E6	223.3E6	55.801	47.097
3)	MA gamma-BHC...	3.900	4.452	609.9E6	209.2E6	56.092	46.091
4)	MA Heptachlor	4.174	4.950	657.5E6	207.5E6	55.880	48.839
5)	MB Aldrin	4.406	5.246	601.9E6	195.7E6	56.965	46.365
6)	B beta-BHC	4.154	4.615	224.3E6	91722532	53.139	45.904
7)	B delta-BHC	4.340	4.823	621.6E6	206.0E6	55.692	44.631
8)	B Heptachlo...	4.836	5.622	571.8E6	176.9E6	55.240	46.073
9)	A Endosulfan I	5.159	5.969	558.6E6	171.6E6	55.415	46.111
10)	B gamma-Chl...	5.054	5.852	612.0E6	186.0E6	56.359	47.074
11)	B alpha-Chl...	5.109	5.925	588.4E6	183.0E6	56.168	46.625
12)	B 4,4'-DDE	5.277	6.083	596.8E6	184.8E6	53.996m	45.890
13)	MA Dieldrin	5.395	6.221	598.3E6	181.0E6	54.619	47.390
14)	MA Endrin	5.643	6.432	549.2E6	158.0E6	55.282m	48.350
15)	B Endosulfa...	5.918	6.636	547.5E6	157.9E6	54.951	45.177
16)	A 4,4'-DDD	5.785	6.561	493.3E6	146.2E6	53.432	46.025
17)	MA 4,4'-DDT	6.013	6.857	479.5E6	140.3E6	59.292m	46.037
18)	B Endrin al...	6.083	6.758	435.7E6	132.5E6	51.893m	44.265
19)	B Endosulfa...	6.294	6.979	504.4E6	141.3E6	52.694	43.935
20)	A Methoxychlor	6.560	7.313	255.0E6	72260775	52.799	47.417
21)	B Endrin ke...	6.769	7.437	545.0E6	144.4E6	50.227	46.984

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100118\
Data File : PL040546.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2018 23:20
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
10/2/2018 8:54:57 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 08:38:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092618.M
Quant Title : GC Extractables
QLast Update : Thu Sep 27 01:27:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

