

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX100323.M
 Title : GC Extractables
 Last Update : Wed Oct 04 05:19:33 2023
 Response Via : Initial Calibration

Calibration Files

500 =PL085667.D 1000 =PL085665.D 250 =PL085668.D
 750 =PL085666.D 100 =PL085669.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.098	3.079	3.023	3.005	2.992	3.039	E6	1.53
2)	Toxaphene-1	1.736	1.766	2.098	1.787	1.650	1.807	E4	9.46
3)	Toxaphene-2	1.777	1.716	1.818	1.738	1.831	1.776	E4	2.79
4)	Toxaphene-3	6.372	6.397	6.254	6.160	6.419	6.320	E4	1.74
5)	Toxaphene-4	3.744	3.733	3.633	3.583	3.654	3.669	E4	1.86
6)	Toxaphene-5	4.553	4.590	4.435	4.454	4.508	4.508	E4	1.44
7) SA	Decachlorobiphenyl	2.480	2.362	2.277	2.323	2.468	2.382	E6	3.75

Signal #2 Calibration Files

500 =PL085667.D 1000 =PL085665.D 250 =PL085668.D
 750 =PL085666.D 100 =PL085669.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.995	9.932	9.516	9.623	9.693	9.752	E5	2.10
2)	Toxaphene-1	5.596	5.536	5.210	5.427	5.714	5.496	E3	3.47
3)	Toxaphene-2	6.374	6.165	5.535	6.168	5.541	5.957	E3	6.57
4)	Toxaphene-3	2.425	2.385	2.219	2.384	2.049	2.293	E4	6.87
5)	Toxaphene-4	3.477	3.409	3.470	3.325	3.550	3.446	E4	2.44
6)	Toxaphene-5	2.288	2.243	2.107	2.237	2.101	2.195	E4	3.89
7) SA	Decachlorobiphenyl	0.951	0.924	0.961	0.923	1.055	0.963	E6	5.63

(#) = Out of Range