

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100419\
 Data File : PL053070.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2019 11:17
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleID :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/7/2019 1:56:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 22:53:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.554	4.049	556.9E6	138.1E6	55.987	51.415
28)	SA Decachlor...	8.347	9.130	435.3E6	131.2E6	43.455	44.715
Target Compounds							
2)	A alpha-BHC	3.998	4.440	909.1E6	200.5E6	56.994	52.208
3)	MA gamma-BHC...	4.285	4.726	829.4E6	183.7E6	56.948	50.736
4)	MA Heptachlor	4.594	5.236	799.5E6	187.7E6	53.507	49.256
5)	MB Aldrin	4.846	5.543	722.9E6	168.9E6	52.555	48.726
6)	B beta-BHC	4.535	4.893	327.1E6	86272131	52.690	54.887m
7)	B delta-BHC	4.744	5.113	770.5E6	177.0E6	53.771	48.907
8)	B Heptachlor...	5.292	5.928	673.1E6	162.2E6	51.651	48.403
9)	A Endosulfan I	5.635	6.286	615.2E6	144.0E6	50.691	46.082
10)	B gamma-Chl...	5.521	6.162	704.3E6	157.9E6	52.373	46.665
11)	B alpha-Chl...	5.580	6.235	673.7E6	155.2E6	51.079	46.123
12)	B 4,4'-DDE	5.744	6.391	595.0E6	152.0E6	51.269	46.524
13)	MA Dieldrin	5.879	6.543	647.8E6	154.1E6	50.189	47.254
14)	MA Endrin	6.138	6.760	588.3E6	130.6E6	51.644	47.046
15)	B Endosulfa...	6.411	6.968	541.6E6	133.4E6	47.262	44.958
16)	A 4,4'-DDD	6.261	6.880	481.5E6	116.7E6	53.572	47.584
17)	MA 4,4'-DDT	6.504	7.179	488.1E6	124.4E6	47.794	47.794
18)	B Endrin al...	6.580	7.092	422.5E6	112.6E6	44.992	45.916
19)	B Endosulfa...	6.793	7.315	470.8E6	123.0E6	45.229	45.040
20)	A Methoxychlor	7.050	7.636	240.8E6	71720956	46.692	50.941
21)	B Endrin ke...	7.287	7.786	512.4E6	126.9E6	43.960	44.479
22)	Mirex	7.482	8.242	386.9E6	102.5E6	40.969	45.563

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100419\
Data File : PL053070.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2019 11:17
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/7/2019 1:56:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 22:53:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

