

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100422\
 Data File : PL078102.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2022 15:32
 Operator : AR\AJ
 Sample : N4960-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PAT-03-100322

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 22:20:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100322.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 03 19:46:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.414	2.623	17326959	32189393	9.395	11.453
28)	SA Decachlor...	8.905	7.721	19908763	27727321	12.682	12.777
Target Compounds							
12)	B 4,4'-DDE	6.073	5.048	8088640	11271329	4.183	3.719m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100422\
Data File : PL078102.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2022 15:32
Operator : AR\AJ
Sample : N4960-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PAT-03-100322

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 22:20:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100322.M
Quant Title : GC Extractables
QLast Update : Mon Oct 03 19:46:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.414 min
Delta R.T.: 0.000 min
Response: 17326959
Conc: 9.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAT-03-100322




#1 Tetrachloro-m-xylene

R.T.: 2.623 min
Delta R.T.: 0.000 min
Response: 32189393
Conc: 11.45 ng/ml




#12 4,4'-DDE

R.T.: 6.073 min
Delta R.T.: -0.002 min
Response: 8088640
Conc: 4.18 ng/ml



#12 4,4'-DDE


R.T.: 5.048 min
Delta R.T.: -0.002 min
Response: 11271329
Conc: 3.72 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.905 min
Delta R.T.: 0.003 min
Response: 19908763
Conc: 12.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAT-03-100322



#28 Decachlorobiphenyl

R.T.: 7.721 min
Delta R.T.: 0.000 min
Response: 27727321
Conc: 12.78 ng/ml