

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100518\
 Data File : PL040702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2018 17:11
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 10/8/2018 10:52:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 05:21:52 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL0100518.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 06 04:55:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.208	3.804	427.7E6	161.6E6	47.669	49.221
28)	SA Decachlor...	7.808	8.728	527.3E6	140.2E6	48.757	49.773
Target Compounds							
2)	A alpha-BHC	3.629	4.176	693.2E6	242.5E6	48.598	50.105
3)	MA gamma-BHC...	3.899	4.450	645.1E6	221.2E6	49.467	49.922
4)	MA Heptachlor	4.173	4.948	697.1E6	210.8E6	48.625	49.553
5)	MB Aldrin	4.404	5.244	621.0E6	194.3E6	50.148	49.486
6)	B beta-BHC	4.153	4.613	235.6E6	93113939	50.189	48.795
7)	B delta-BHC	4.339	4.821	650.9E6	212.2E6	49.806	51.799
8)	B Heptachlo...	4.834	5.620	592.7E6	174.7E6	49.235	51.602
9)	A Endosulfan I	5.158	5.967	574.7E6	165.6E6	48.986	49.011
10)	B gamma-Chl...	5.053	5.850	631.3E6	177.3E6	48.955	50.623
11)	B alpha-Chl...	5.107	5.923	608.7E6	175.9E6	48.967	49.007
12)	B 4,4'-DDE	5.277	6.082	627.1E6	175.1E6	49.106	48.726
13)	MA Dieldrin	5.394	6.219	629.0E6	175.7E6	49.238	49.183
14)	MA Endrin	5.643	6.431	582.4E6	157.2E6	48.304	49.109
15)	B Endosulfa...	5.917	6.635	567.9E6	158.4E6	49.412	45.829
16)	A 4,4'-DDD	5.783	6.560	505.9E6	138.4E6	52.651	50.268
17)	MA 4,4'-DDT	6.013	6.855	518.8E6	145.1E6	50.525	47.529
18)	B Endrin al...	6.083	6.756	460.9E6	130.2E6	47.891	48.653
19)	B Endosulfa...	6.293	6.977	530.0E6	143.0E6	48.777	49.021
20)	A Methoxychlor	6.559	7.312	277.3E6	77588724	49.647	48.724
21)	B Endrin ke...	6.767	7.435	599.8E6	151.3E6	49.658m	49.544

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100518\
Data File : PL040702.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2018 17:11
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
10/8/2018 10:52:57 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 05:21:52 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL0100518.M
Quant Title : GC Extractables
QLast Update : Sat Oct 06 04:55:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

