

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100722\
 Data File : PL078185.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 15:13
 Operator : AR\AJ
 Sample : N4519-02
 Misc : TOX MDL 100PPB
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MDL-WATER-QT4-2022-08

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/07/2022
 Supervised By :Ankita Jodhani 10/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 00:35:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100522CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 01:39:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.419	2.620	32064240	57313258	18.178	17.035
27)	SA Decachlor...	8.911	7.721	62641915	99578394	41.328	36.775
Target Compounds							
22)	Toxaphene-1	6.122	5.138	1731667	1962554	123.740m	121.346m
23)	Toxaphene-2	6.945	6.404	4258170	6294935	104.128m	93.610m
24)	Toxaphene-3	7.036	6.530	3525038	9909921	110.019m	108.826m
25)	Toxaphene-4	7.127	6.843	1970701	6693602	92.658m	122.078 #
26)	Toxaphene-5	7.824	6.975	2606350	2652351	108.488m	80.470m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100722\
Data File : PL078185.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 15:13
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100PPB
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 00:35:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100522CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 05 01:39:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 10/07/2022
Supervised By: Ankita Jodhani 10/07/2022

Manual Integrations
APPROVED

Instrument :
ECD_L
Client Sample :
MDL-WATER-QT4-2022-08