

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100818\
 Data File : PL040775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2018 15:40
 Operator : AJ\SJ
 Sample : PB113592BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 01:07:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL0100518.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 06 04:55:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.208	3.804	158.4E6	60460382	17.656	18.419
28) SA Decachlor...	7.806	8.725	131.7E6	51627773	12.177	18.334 #
Target Compounds						
4) MA Heptachlor	4.107f	0.000	5075124	0	0.354	N.D. #
6) B beta-BHC	4.107f	0.000	5075124	0	1.081	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100818\
Data File : PL040775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2018 15:40
Operator : AJ\SJ
Sample : PB113592BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 01:07:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL0100518.M
Quant Title : GC Extractables
QLast Update : Sat Oct 06 04:55:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.208 min
Delta R.T.: 0.000 min
Response: 158405059
Conc: 17.66 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.804 min
Delta R.T.: 0.000 min
Response: 60460382
Conc: 18.42 ng/ml




#4 Heptachlor

R.T.: 4.107 min
Delta R.T.: -0.065 min
Response: 5075124
Conc: 0.35 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T. : 4.947 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.107 min
Delta R.T.: -0.045 min
Response: 5075124
Conc: 1.08 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.612 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.806 min
Delta R.T.: -0.003 min
Response: 131684359
Conc: 12.18 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.725 min
Delta R.T.: -0.002 min
Response: 51627773
Conc: 18.33 ng/ml