

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100819\
 Data File : PL053154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2019 17:13
 Operator : SG\AJ
 Sample : K5238-01MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-COMPMS

Manual Integrations
 APPROVED

Ankita
 10/9/2019 12:36:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 00:41:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	4.047	178.0E6	49126783	17.893	18.295
28)	SA Decachlor...	8.336	9.125	108.1E6	34923112	10.792	11.904
Target Compounds							
2)	A alpha-BHC	3.990	4.439	749.3E6	171.7E6	46.974	44.718
3)	MA gamma-BHC...	4.277	4.724	710.6E6	163.5E6	48.791	45.147
4)	MA Heptachlor	4.585	5.234	749.4E6	171.8E6	50.151	45.085
5)	MB Aldrin	4.837	5.540	685.0E6	152.9E6	49.799	44.112
6)	B beta-BHC	4.527	4.894	304.8E6	76112186	49.089	48.423
7)	B delta-BHC	4.736	5.111	465.9E6	105.3E6	32.515	29.102
8)	B Heptachlor...	5.283	5.926	614.6E6	144.9E6	47.163	43.243
9)	A Endosulfan I	5.625	6.282	561.5E6	129.5E6	46.270	41.427
10)	B gamma-Chl...	5.512	6.159	623.6E6	139.3E6	46.369	41.165
11)	B alpha-Chl...	5.571	6.232	603.5E6	137.3E6	45.758	40.778
12)	B 4,4'-DDE	5.734	6.388	569.8E6	133.1E6	49.094	40.759
13)	MA Dieldrin	5.870	6.539	593.0E6	128.8E6	45.938	39.475
14)	MA Endrin	6.129	6.757	510.7E6	114.6E6	44.834	41.277
15)	B Endosulfa...	6.402	6.965	462.0E6	117.1E6	40.311	39.471
16)	A 4,4'-DDD	6.251	6.877	391.0E6	99265203	43.501	40.474
17)	MA 4,4'-DDT	6.494	7.176	429.8E6	102.9E6	42.089	39.527
18)	B Endrin al...	6.570	7.089	366.4E6	94123547	39.014	38.376
19)	B Endosulfa...	6.785	7.312	378.6E6	98683754	36.372	36.142
20)	A Methoxychlor	7.040	7.633	205.9E6	57524824	39.926	40.858
21)	B Endrin ke...	7.278	7.783	435.7E6	114.6E6	37.384	40.183
22)	Mirex	7.473	8.237	383.0E6	99043576	40.553	44.034m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100819\
Data File : PL053154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2019 17:13
Operator : SG\AJ
Sample : K5238-01MS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
TP-COMPMS

Manual Integrations
APPROVED

Ankita
10/9/2019 12:36:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 00:41:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

