

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL100821\  
 Data File : PL070208.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 08 Oct 2021 17:38  
 Operator : AR\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_L  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Oct 09 00:10:33 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL092821.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Sep 29 09:19:28 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

|                             | Compound        | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|-----------------|--------|--------|----------|----------|--------|--------|
| -----                       |                 |        |        |          |          |        |        |
| System Monitoring Compounds |                 |        |        |          |          |        |        |
| 1)                          | SA Tetrachlo... | 4.625  | 5.562  | 9925384  | 30433477 | 19.835 | 23.010 |
| 28)                         | SA Decachlor... | 10.390 | 11.481 | 14541436 | 27809659 | 21.014 | 22.980 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL100821\  
Data File : PL070208.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 08 Oct 2021 17:38  
Operator : AR\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 09 00:10:33 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL092821.M  
Quant Title : GC Extractables  
QLast Update : Wed Sep 29 09:19:28 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.625 min  
Delta R.T.: -0.004 min  
Response: 9925384  
Conc: 19.84 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 5.562 min  
Delta R.T.: -0.005 min  
Response: 30433477  
Conc: 23.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.390 min  
Delta R.T.: -0.009 min  
Response: 14541436  
Conc: 21.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.481 min  
Delta R.T.: -0.011 min  
Response: 27809659  
Conc: 22.98 ng/ml