

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX100918.M
 Title : GC Extractables
 Last Update : Wed Oct 10 02:07:40 2018
 Response Via : Initial Calibration

Calibration Files

500 =PL040821.D 1000 =PL040819.D 250 =PL040822.D
 750 =PL040820.D 50 =PL040823.D

Compound		500	1000	250	750	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.845	4.874	4.647	4.802	4.286	4.691	E6	5.17
2)	Toxaphene-1	3.763	3.736	4.025	3.698	3.469	3.738	E4	5.30
3)	Toxaphene-2	9.110	9.317	9.096	9.138	7.302	8.793	E4	9.53
4)	Toxaphene-3	2.786	2.672	2.371	2.765	2.081	2.535	E4	11.97
5)	Toxaphene-4	4.727	4.875	4.384	4.889	3.833	4.542	E4	9.80
6)	Toxaphene-5	8.765	8.803	8.972	8.712	7.562	8.563	E4	6.63
7) SA	Decachlorobiphenyl	4.643	4.481	4.714	4.514	4.780	4.626	E6	2.76

Signal #2 Calibration Files

500 =PL040821.D 1000 =PL040819.D 250 =PL040822.D
 750 =PL040820.D 50 =PL040823.D

Compound		500	1000	250	750	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.734	2.700	2.617	2.693	2.578	2.664	E6	2.42
2)	Toxaphene-1	7.209	6.835	6.752	6.661	6.696	6.831	E3	3.24
3)	Toxaphene-2	1.642	1.663	1.710	1.630	1.742	1.677	E4	2.80
4)	Toxaphene-3	6.874	6.858	6.900	6.664	7.046	6.868	E4	1.98
5)	Toxaphene-4	4.139	4.184	4.215	4.093	3.752	4.077	E4	4.59
6)	Toxaphene-5	1.838	1.841	1.931	1.765	1.728	1.821	E4	4.31
7) SA	Decachlorobiphenyl	2.484	2.335	2.461	2.380	2.497	2.431	E6	2.90

(#) = Out of Range