

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100919\
 Data File : PL053189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2019 17:05
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/10/2019 10:41:40 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 00:44:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	4.048	547.0E6	128.8E6	54.991	47.950
28)	SA Decachlor...	8.341	9.128	474.5E6	132.7E6	47.371	45.228
Target Compounds							
2)	A alpha-BHC	3.995	4.440	900.7E6	186.4E6	56.466	48.534
3)	MA gamma-BHC...	4.282	4.723	821.6E6	175.3E6	56.408	48.408m
4)	MA Heptachlor	4.590	5.235	823.4E6	178.0E6	55.106	46.708
5)	MB Aldrin	4.843	5.541	768.7E6	161.8E6	55.885	46.677
6)	B beta-BHC	4.533	4.892	329.2E6	85344573	53.019	54.297m
7)	B delta-BHC	4.741	5.112	806.9E6	170.6E6	56.308	47.131
8)	B Heptachlor...	5.289	5.927	709.5E6	157.6E6	54.449	47.028
9)	A Endosulfan I	5.630	6.284	657.6E6	144.8E6	54.190	46.332
10)	B gamma-Chl...	5.517	6.161	729.4E6	158.0E6	54.237	46.704
11)	B alpha-Chl...	5.576	6.234	709.9E6	155.2E6	53.826	46.119
12)	B 4,4'-DDE	5.739	6.389	642.6E6	152.3E6	55.363	46.618
13)	MA Dieldrin	5.876	6.541	711.2E6	151.5E6	55.094	46.459
14)	MA Endrin	6.134	6.759	628.2E6	130.5E6	55.150	46.992
15)	B Endosulfa...	6.407	6.967	597.7E6	138.7E6	52.156	46.734
16)	A 4,4'-DDD	6.257	6.879	500.4E6	117.4E6	55.670	47.884
17)	MA 4,4'-DDT	6.500	7.178	550.5E6	123.8E6	53.903	47.564
18)	B Endrin al...	6.575	7.091	484.2E6	113.9E6	51.557	46.450
19)	B Endosulfa...	6.789	7.314	533.4E6	126.3E6	51.240	46.256
20)	A Methoxychlor	7.046	7.635	269.7E6	65751194	52.287	46.701
21)	B Endrin ke...	7.283	7.785	584.1E6	132.8E6	50.113	46.545
22)	Mirex	7.478	8.241	438.6E6	98870763	46.450	43.957

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100919\
Data File : PL053189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2019 17:05
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/10/2019 10:41:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 00:44:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

