

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100923\
 Data File : PL085767.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2023 12:52
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 03:13:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.362	2.623	145.7E6	45350162	48.999	48.549
28)	SA Decachlor...	8.864	7.777	121.5E6	49013795	55.768	53.104
Target Compounds							
2)	A alpha-BHC	3.816	3.123	214.1E6	61449833	49.893	48.842
3)	MA gamma-BHC...	4.148	3.450	206.3E6	60438108	49.908	49.003
4)	MA Heptachlor	4.736	3.787	197.7E6	56116549	51.115	48.249
5)	MB Aldrin	5.075	4.064	194.0E6	54918639	50.058	49.111
6)	B beta-BHC	4.350	3.752	85126637	26090847	51.479	50.118
7)	B delta-BHC	4.593	3.977	200.5E6	58460452	52.279	50.881
8)	B Heptachlo...	5.505	4.568	173.9E6	49644627	50.419	49.427
9)	A Endosulfan I	5.889	4.936	161.4E6	45565824	49.900	42.476
10)	B gamma-Chl...	5.762	4.820	175.8E6	48747183	51.376	49.220
11)	B alpha-Chl...	5.842	4.884	173.8E6	49799428	50.863	45.657
12)	B 4,4'-DDE	6.025	5.083	150.4E6	44012528	53.781	53.600
13)	MA Dieldrin	6.165	5.202	165.0E6	48391341	50.206	49.877
14)	MA Endrin	6.394	5.476	146.9E6	45395625	51.537	50.495
15)	B Endosulfa...	6.617	5.774	140.4E6	46096889	50.953	48.475
16)	A 4,4'-DDD	6.543	5.640	119.6E6	38521341	59.362	55.343
17)	MA 4,4'-DDT	6.859	5.890	128.6E6	43319719	54.540	54.783
18)	B Endrin al...	6.747	5.955	105.0E6	36770764	51.807	49.310
19)	B Endosulfa...	6.984	6.180	134.1E6	46472922	53.420	51.515
20)	A Methoxychlor	7.344	6.475	71883287	26725918	51.469	50.802
21)	B Endrin ke...	7.465	6.682	143.8E6	57911400	49.927	50.383
22)	Mirex	7.938	6.859	111.2E6	47138319	50.248	49.979

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL100923\
Data File : PL085767.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2023 12:52
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 03:13:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
Quant Title : GC Extractables
QLast Update : Wed Oct 04 05:29:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

