

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101019\
 Data File : PL053204.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2019 13:11
 Operator : SG\AJ
 Sample : MDL-W-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleID :
 MDL-W-03

Manual Integrations
 APPROVED

Ankita
 10/11/2019 2:07:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 08:14:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092819CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 28 05:45:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	4.047	162.9E6	41089907	17.505	16.362
27)	SA Decachlor...	8.336	9.125	293.5E6	86008472	32.373	32.200
Target Compounds							
2)	A alpha-BHC	3.989	4.438	12523768	2497693	0.888	0.741
3)	MA gamma-BHC...	4.275	4.723	9638936	2497412	0.746m	0.773
4)	MA Heptachlor	4.584	5.233	11617603	3437181	0.851	0.987
5)	MB Aldrin	4.836	5.539	9556527	2713459	0.772	0.872
6)	B beta-BHC	4.526	4.894	5925238	1836252	1.009	1.237
7)	B delta-BHC	4.735	5.108	8188227	4124871	0.634	1.237 #
8)	B Heptachlo...	5.282	5.927	9875237	4806627	0.830	1.553 #
9)	A Endosulfan I	5.623	6.282	9111326	2486143	0.826m	0.866
10)	B trans-Chl...	5.511	6.159	9861516	2543868	0.815	0.838
11)	B cis-Chlor...	5.570	6.232	9775431	2550235	0.833	0.855
12)	B 4,4'-DDE	5.733	6.387	7946059	2781739	0.733	0.923 #
13)	MA Dieldrin	5.869	6.539	8879645	2579206	0.724	0.843
14)	MA Endrin	6.128	6.755	9209730	2066655	0.918	0.854m
15)	B Endosulfa...	6.399	6.966	9583715	3003132	0.901m	1.115
16)	A 4,4'-DDD	6.249	6.876	6161077	2770375	0.749m	1.246 #
17)	MA 4,4'-DDT	6.494	7.176	6691098	1935318	0.701	0.817
18)	B Endrin al...	6.569	7.089	7710575	2107677	0.914	0.961
19)	B Endosulfa...	6.783	7.312	8616428	2198738	0.880	0.879
20)	A Methoxychlor	7.040	7.632	4249691	1093277	0.900	0.848
21)	B Endrin ke...	7.277	7.783	9139928	2289395	0.863	0.899

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101019\
Data File : PL053204.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2019 13:11
Operator : SG\AJ
Sample : MDL-W-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
Client Sample Id :
MDL-W-03

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 08:14:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092819CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 28 05:45:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Manual Integrations
APPROVED

Ankita
10/11/2019 2:07:09 PM

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

