

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101118\
 Data File : PL040937.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2018 19:35
 Operator : AJ\SJ
 Sample : J5367-01
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 DS-VB-16202

Manual Integrations
 APPROVED

Sohil
 10/12/2018 5:21:37 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 01:39:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101118.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 11 03:28:36 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.480	4.105	84575677	38914966	15.431	13.957
28)	SA Decachlor...	8.233	9.240	31090794	27739619	5.980m	10.727 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101118\
Data File : PL040937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2018 19:35
Operator : AJ\SJ
Sample : J5367-01
Misc :
ALS Vial : 39 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
DS-VB-16202


Manual Integrations
APPROVED

Sohil
10/12/2018 5:21:37 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 01:39:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101118.M
Quant Title : GC Extractables
QLast Update : Thu Oct 11 03:28:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: 0.000 min
Response: 84575677
Conc: 15.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
DS-VB-16202


Manual Integrations
APPROVED

Sohil
10/12/2018 5:21:37 PM




#1 Tetrachloro-m-xylene

R.T.: 4.105 min
Delta R.T.: 0.002 min
Response: 38914966
Conc: 13.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.233 min
Delta R.T.: -0.001 min
Response: 31090794
Conc: 5.98 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.240 min
Delta R.T.: 0.009 min
Response: 27739619
Conc: 10.73 ng/ml