

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101120\
 Data File : PL062197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2020 21:17
 Operator : DD\AJ
 Sample : INDA352
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA352

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 01:47:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100920CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 10 05:27:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.952	32787969	52528426	18.388	18.431
27)	SA Decachlor...	8.124	8.992	69742079	103.7E6	36.688	37.314
Target Compounds							
2)	A alpha-BHC	3.844	4.338	43553461	71723530	17.831	17.925
3)	MA gamma-BHC...	4.123	4.621	38732450	67598651	17.118	18.127
4)	MA Heptachlor	4.420	5.127	37725220	69055346	18.736	18.488
9)	A Endosulfan I	5.437	6.171	30271536	53575999	18.710	18.086
13)	MA Dieldrin	5.678	6.428	65743830	113.7E6	36.860	37.078
14)	MA Endrin	5.933	6.644	56630702	89879214	35.483	36.891
16)	A 4,4'-DDD	6.061	6.769	52166113	87375586	37.351	37.285
17)	MA 4,4'-DDT	6.300	7.067	49046966	83476660	37.072	36.671
20)	A Methoxychlor	6.844	7.527	153.6E6	243.9E6	187.653	183.408

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101120\
Data File : PL062197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2020 21:17
Operator : DD\AJ
Sample : INDA352
Misc :
ALS Vial : 51 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 01:47:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100920CLP.M
Quant Title : GC Extractables
QLast Update : Sat Oct 10 05:27:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

