

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101220\
 Data File : PL062218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2020 12:49
 Operator : DD\AJ
 Sample : MDL-TOX-13
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MDL-TOX-13

Manual Integrations
 APPROVED

Ankita
 10/13/2020 12:27:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 13:12:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100920CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 10 05:27:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.951	26056427	41724570	14.613	14.640
27)	SA Decachlor...	8.124	8.991	55918559	84634189	29.416	30.453
Target Compounds							
22)	Toxaphene-1	5.641	6.328	860439	1154150	81.114	84.582m
23)	Toxaphene-2	5.800	6.618	575044	1494115	64.579	81.683 #
24)	Toxaphene-3	6.237	7.024	2070266	1581452	72.587	66.893
25)	Toxaphene-4	6.481	7.279	1992007	1185095	66.471	64.810
26)	Toxaphene-5	6.964	7.600	2092721	2448401	67.476	77.099

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101220\
Data File : PL062218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2020 12:49
Operator : DD\AJ
Sample : MDL-TOX-13
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 13:12:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100920CLP.M
Quant Title : GC Extractables
QLast Update : Sat Oct 10 05:27:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
10/13/2020 12:27:35 PM

Instrument :
ECD_L
Client/SampleId :
MDL-TOX-13