

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101320\
 Data File : PL062232.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2020 18:42
 Operator : DD\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 03:38:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 13 03:38:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.952	168.6E6	277.9E6	97.205	97.135
28)	SA Decachlor...	8.125	8.992	167.6E6	252.0E6	92.713	92.516
Target Compounds							
2)	A alpha-BHC	3.845	4.338	260.8E6	440.4E6	101.434	101.146
3)	MA gamma-BHC...	4.124	4.622	232.7E6	405.3E6	101.228	100.011
4)	MA Heptachlor	4.421	5.128	210.0E6	380.3E6	99.102	98.915
5)	MB Aldrin	4.665	5.431	191.9E6	363.4E6	100.434	100.588
6)	B beta-BHC	4.372	4.793	91320690	165.3E6	95.315	94.900
7)	B delta-BHC	4.573	5.008	219.0E6	361.0E6	100.416	100.382
8)	B Heptachlo...	5.103	5.817	179.6E6	323.6E6	97.890	99.036
9)	A Endosulfan I	5.438	6.173	166.5E6	299.1E6	98.282	98.276
10)	B gamma-Chl...	5.328	6.051	179.0E6	329.6E6	99.076	99.071
11)	B alpha-Chl...	5.386	6.123	178.4E6	324.7E6	98.514	98.135
12)	B 4,4'-DDE	5.551	6.282	169.0E6	310.8E6	101.449	100.527
13)	MA Dieldrin	5.679	6.429	177.8E6	314.0E6	100.217	100.059
14)	MA Endrin	5.933	6.645	153.8E6	251.2E6	99.698	100.193
15)	B Endosulfa...	6.205	6.855	161.9E6	265.7E6	97.437	97.345
16)	A 4,4'-DDD	6.062	6.771	147.3E6	249.2E6	100.295	99.753
17)	MA 4,4'-DDT	6.300	7.069	140.8E6	240.1E6	103.870	102.144
18)	B Endrin al...	6.372	6.979	139.0E6	227.3E6	96.104	96.120
19)	B Endosulfa...	6.584	7.202	158.4E6	251.0E6	95.992	95.982
20)	A Methoxychlor	6.844	7.528	81010623	128.7E6	96.385	96.055
21)	B Endrin ke...	7.069	7.671	192.1E6	274.4E6	96.609	96.807
22)	Mirex	7.256	8.124	136.3E6	192.5E6	92.660	92.286

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101320\
Data File : PL062232.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2020 18:42
Operator : DD\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 03:38:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
Quant Title : GC Extractables
QLast Update : Tue Oct 13 03:38:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

