

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101320\
 Data File : PL062239.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2020 20:22
 Operator : DD\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 04:20:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 13 04:19:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.952	99052206	140.1E6	50.000	50.000
28)	SA Decachlor...	8.125	8.993	90135881	135.2E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.270	4.940	26296224	51121697	500.000	500.000
24)	Chlordane-2	4.766	5.406	30170086	53113930	500.000	500.000
25)	Chlordane-3	5.328	6.051	81435952	210.1E6	500.000	500.000
26)	Chlordane-4	5.385	6.125	68017180	253.2E6	500.000	500.000
27)	Chlordane-5	6.213	6.926	27280796	41398746	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101320\
Data File : PL062239.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2020 20:22
Operator : DD\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 04:20:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
Quant Title : GC Extractables
QLast Update : Tue Oct 13 04:19:56 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.412 min
Delta R.T.: 0.000 min
Response: 99052206
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.000 min
Response: 140062806
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.270 min
Delta R.T.: 0.000 min
Response: 26296224
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 4.940 min
Delta R.T.: 0.000 min
Response: 51121697
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 30170086
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 5.406 min
Delta R.T.: 0.000 min
Response: 53113930
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 5.328 min
Delta R.T.: 0.000 min
Response: 81435952
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 6.051 min
Delta R.T.: 0.000 min
Response: 210119256
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 5.385 min
Delta R.T.: 0.000 min
Response: 68017180
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#26 Chlordane-4

R.T.: 6.125 min
Delta R.T.: 0.000 min
Response: 253211908
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 6.213 min
Delta R.T.: 0.000 min
Response: 27280796
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 6.926 min
Delta R.T.: 0.000 min
Response: 41398746
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.125 min
Delta R.T.: 0.000 min
Response: 90135881
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 8.993 min
Delta R.T.: 0.000 min
Response: 135239183
Conc: 50.00 ng/ml