

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101518\
 Data File : PL040996.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2018 14:22
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 00:32:13 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101118.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 11 03:28:36 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.478	4.104	268.3E6	137.6E6	48.955	49.361
28)	SA Decachlor...	8.228	9.233	243.0E6	113.1E6	46.739	43.753
Target Compounds							
2)	A alpha-BHC	3.919	4.499	404.0E6	195.0E6	50.991	50.398
3)	MA gamma-BHC...	4.203	4.788	369.1E6	180.1E6	50.545	50.582
4)	MA Heptachlor	4.501	5.304	363.0E6	168.5E6	49.257	49.775
5)	MB Aldrin	4.748	5.613	344.4E6	153.4E6	50.175	50.832
6)	B beta-BHC	4.455	4.956	155.3E6	81164919	48.936	50.642
7)	B delta-BHC	4.659	5.177	344.2E6	168.1E6	51.655	51.223
8)	B Heptachlo...	5.192	6.001	319.5E6	141.4E6	48.715	49.305
9)	A Endosulfan I	5.530	6.361	293.0E6	124.5E6	48.906	48.621
10)	B gamma-Chl...	5.417	6.235	322.3E6	139.1E6	49.002m	49.589
11)	B alpha-Chl...	5.476	6.309	314.4E6	135.7E6	48.323	48.873
12)	B 4,4'-DDE	5.642	6.463	285.0E6	128.3E6	50.144	48.510
13)	MA Dieldrin	5.773	6.619	314.8E6	125.4E6	49.797	47.098
14)	MA Endrin	6.030	6.839	269.8E6	101.2E6	46.964	45.333
15)	B Endosulfa...	6.304	7.046	265.9E6	112.7E6	47.441	47.437
16)	A 4,4'-DDD	6.159	6.955	235.0E6	102.0E6	52.034	48.671
17)	MA 4,4'-DDT	6.398	7.255	204.0E6	89923502	44.791	43.230
18)	B Endrin al...	6.472	7.170	207.2E6	92917543	49.406m	46.831
19)	B Endosulfa...	6.686	7.394	233.9E6	105.4E6	45.973	47.176
20)	A Methoxychlor	6.946	7.710	111.2E6	49445031	42.765	42.004
21)	B Endrin ke...	7.176	7.868	274.4E6	112.1E6	52.413	47.650

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101518\
Data File : PL040996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2018 14:22
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 00:32:13 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101118.M
Quant Title : GC Extractables
QLast Update : Thu Oct 11 03:28:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

