

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101519\
 Data File : PL053325.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2019 19:51
 Operator : SG\AJ
 Sample : PB123934BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB123934BS

Manual Integrations
 APPROVED

Ankita
 10/15/2019 11:38:06 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 00:56:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	4.046	197.2E6	50894265	19.823	18.954
28)	SA Decachlor...	8.335	9.125	188.1E6	58099288	18.779	19.804
Target Compounds							
2)	A alpha-BHC	3.989	4.438	847.3E6	187.4E6	53.121	48.799
3)	MA gamma-BHC...	4.276	4.723	771.1E6	176.7E6	52.943	48.793
4)	MA Heptachlor	4.583	5.233	792.4E6	180.2E6	53.031	47.279
5)	MB Aldrin	4.835	5.538	767.9E6	173.2E6	55.824	49.989
6)	B beta-BHC	4.526	4.892	318.6E6	81482310	51.309	51.840
7)	B delta-BHC	4.735	5.110	719.6E6	164.4E6	50.221	45.422
8)	B Heptachlor...	5.282	5.924	688.3E6	167.8E6	52.818	50.061
9)	A Endosulfan I	5.624	6.282	620.1E6	153.2E6	51.101	49.009
10)	B gamma-Chl...	5.510	6.158	692.9E6	166.5E6	51.521	49.216
11)	B alpha-Chl...	5.569	6.231	664.1E6	163.2E6	50.351	48.498
12)	B 4,4'-DDE	5.732	6.387	534.9E6	160.4E6	46.088	49.125
13)	MA Dieldrin	5.869	6.538	671.2E6	162.6E6	51.997	49.855
14)	MA Endrin	6.127	6.756	564.4E6	141.1E6	49.552	50.826
15)	B Endosulfa...	6.400	6.965	546.2E6	149.7E6	47.665	50.458
16)	A 4,4'-DDD	6.250	6.875	480.2E6	125.2E6	53.427	51.029m
17)	MA 4,4'-DDT	6.493	7.176	487.7E6	125.4E6	47.754	48.200
18)	B Endrin al...	6.569	7.088	452.7E6	126.1E6	48.210	51.430
19)	B Endosulfa...	6.782	7.311	524.7E6	137.8E6	50.407	50.456
20)	A Methoxychlor	7.037	7.633	236.6E6	71973778	45.863m	51.120
21)	B Endrin ke...	7.275	7.783	642.6E6	149.3E6	55.137m	52.323
22)	Mirex	7.470	8.238	428.3E6	111.8E6	45.354	49.727

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101519\
Data File : PL053325.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2019 19:51
Operator : SG\AJ
Sample : PB123934BS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB123934BS

Manual Integrations
APPROVED

Ankita
10/15/2019 11:38:06 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 00:56:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

