

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101524\
 Data File : PL092387.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2024 19:13
 Operator : AR\AJ
 Sample : P4401-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SU-03-101424

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/16/2024
 Supervised By : Ankita Jodhani 10/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 03:44:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092324.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 23 16:44:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.780	34652034	38050749	15.576	15.171
28) SA Decachlor...	9.060	7.921	25508263	31608511	15.571	12.634
Target Compounds						
17) MA 4,4'-DDT	7.025	6.041	1195528	1791123	0.649m	0.774m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101524\
Data File : PL092387.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2024 19:13
Operator : AR\AJ
Sample : P4401-01
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
SU-03-101424


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/16/2024
Supervised By : Ankita Jodhani 10/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 03:44:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092324.M
Quant Title : GC Extractables
QLast Update : Mon Sep 23 16:44:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: -0.003 min
Response: 34652034
Conc: 15.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
SU-03-101424


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/16/2024
Supervised By :Ankita Jodhani 10/16/2024




#1 Tetrachloro-m-xylene

R.T.: 2.780 min
Delta R.T.: -0.002 min
Response: 38050749
Conc: 15.17 ng/ml




#17 4,4'-DDT

R.T.: 7.025 min
Delta R.T.: -0.003 min
Response: 1195528
Conc: 0.65 ng/ml m



#17 4,4'-DDT

R.T.: 6.041 min
Delta R.T.: -0.005 min
Response: 1791123
Conc: 0.77 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: 0.000 min
Response: 25508263
Conc: 15.57 ng/ml

Instrument :

ECD_L

ClientSampleId :


SU-03-101424

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/16/2024

Supervised By :Ankita Jodhani 10/16/2024



#28 Decachlorobiphenyl

R.T.: 7.921 min
Delta R.T.: -0.002 min
Response: 31608511
Conc: 12.63 ng/ml