

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101619\
 Data File : PL053352.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2019 17:53
 Operator : SG\AJ
 Sample : K5355-08
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 00:45:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 30 08:44:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	4.046	182.0E6	42233763	18.294	15.728
28)	SA Decachlor...	8.334	9.125	180.6E6	52558467	18.033	17.915
Target Compounds							
7)	B delta-BHC	0.000	5.085f	0	3116769	N.D.	0.861 #
8)	B Heptachlo...	0.000	5.912	0	606273	N.D.	0.181 #
13)	MA Dieldrin	0.000	6.511f	0	1121347	N.D.	0.344 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101619\
Data File : PL053352.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2019 17:53
Operator : SG\AJ
Sample : K5355-08
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 00:45:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092919.M
Quant Title : GC Extractables
QLast Update : Mon Sep 30 08:44:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: -0.004 min
Response: 181960263
Conc: 18.29 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.046 min
Delta R.T.: -0.001 min
Response: 42233763
Conc: 15.73 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.740 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.085 min
Delta R.T.: -0.024 min
Response: 3116769
Conc: 0.86 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 5.289 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.912 min
Delta R.T.: -0.013 min
Response: 606273
Conc: 0.18 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 5.876 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 6.511 min
Delta R.T.: -0.027 min
Response: 1121347
Conc: 0.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.334 min
Delta R.T.: -0.011 min
Response: 180624725
Conc: 18.03 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.125 min
Delta R.T.: 0.000 min
Response: 52558467
Conc: 17.92 ng/ml