

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101620\
 Data File : PL062362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2020 21:11
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita

10/16/2020 11:52:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 05:33:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 13 05:04:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.952	85982755	139.3E6	49.399	49.434
28)	SA Decachlor...	8.125	8.993	84835409	121.9E6	46.377	44.402
Target Compounds							
2)	A alpha-BHC	3.844	4.339	126.7E6	211.0E6	49.431	50.323
3)	MA gamma-BHC...	4.123	4.622	112.7E6	195.7E6	49.438	49.814
4)	MA Heptachlor	4.420	5.127	102.0E6	182.4E6	48.466	47.991
5)	MB Aldrin	4.664	5.430	94213788	176.5E6	49.697	50.056
6)	B beta-BHC	4.372	4.793	47619683	84208298	49.514	47.920
7)	B delta-BHC	4.573	5.007	106.3E6	180.1E6	49.828	51.920
8)	B Heptachlo...	5.102	5.816	90167046	165.7E6	49.340m	50.974
9)	A Endosulfan I	5.438	6.172	83651242	147.8E6	49.544	49.062
10)	B gamma-Chl...	5.327	6.050	89334661	162.5E6	49.545m	49.475
11)	B alpha-Chl...	5.385	6.123	89332741	160.8E6	49.352	49.194
12)	B 4,4'-DDE	5.550	6.282	81987005	150.5E6	49.719	49.462
13)	MA Dieldrin	5.678	6.429	87012934	152.6E6	49.500	49.405
14)	MA Endrin	5.933	6.646	71446664	112.1E6	46.505	45.070
15)	B Endosulfa...	6.204	6.855	80897257	128.9E6	48.995	47.494
16)	A 4,4'-DDD	6.061	6.771	76610583	125.5E6	52.821	51.428
17)	MA 4,4'-DDT	6.301	7.069	58179129	97969879	45.101	42.831
18)	B Endrin al...	6.372	6.979	60307414	95333880	41.561	40.358
19)	B Endosulfa...	6.584	7.202	79872429	122.3E6	48.321	47.115
20)	A Methoxychlor	6.844	7.529	36098048	54850458	44.360	41.610
21)	B Endrin ke...	7.068	7.671	97887666	134.1E6	49.682m	48.097
22)	Mirex	7.256	8.124	70742332	94107379	47.671	44.721

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101620\
Data File : PL062362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2020 21:11
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita

10/16/2020 11:52:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 05:33:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
Quant Title : GC Extractables
QLast Update : Tue Oct 13 05:04:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

