

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101720\
 Data File : PL062401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2020 17:33
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/19/2020 3:06:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 02:37:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 13 05:04:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.951	95923510	153.6E6	55.110	54.483
28)	SA Decachlor...	8.123	8.991	85683650	122.9E6	46.841	44.771
Target Compounds							
2)	A alpha-BHC	3.844	4.337	142.0E6	236.6E6	55.402	56.410
3)	MA gamma-BHC...	4.123	4.620	126.9E6	218.3E6	55.672	55.565
4)	MA Heptachlor	4.419	5.125	114.0E6	203.7E6	54.165	53.594
5)	MB Aldrin	4.664	5.429	105.6E6	195.4E6	55.695	55.440
6)	B beta-BHC	4.371	4.791	52789090	92903112	54.889	52.868
7)	B delta-BHC	4.572	5.006	120.4E6	198.8E6	56.450	57.338
8)	B Heptachlo...	5.102	5.815	100.1E6	187.6E6	54.768	57.725
9)	A Endosulfan I	5.436	6.170	92711207	156.2E6	54.910	51.856
10)	B gamma-Chl...	5.326	6.049	99041133	173.0E6	54.928m	52.672
11)	B alpha-Chl...	5.384	6.121	99137103	170.2E6	54.769	52.080
12)	B 4,4'-DDE	5.549	6.280	91660151	158.1E6	55.586	51.975
13)	MA Dieldrin	5.677	6.427	97074106	160.3E6	55.224	51.909
14)	MA Endrin	5.932	6.644	83567241	126.1E6	54.394	50.711
15)	B Endosulfa...	6.204	6.853	87978371	133.9E6	53.284	49.336
16)	A 4,4'-DDD	6.061	6.769	86480500	134.3E6	59.626	55.013
17)	MA 4,4'-DDT	6.299	7.067	59586614	96027278	46.192	41.981
18)	B Endrin al...	6.370	6.977	67190173	101.9E6	46.304	43.150
19)	B Endosulfa...	6.583	7.200	85899945	125.4E6	51.968	48.313
20)	A Methoxychlor	6.843	7.526	38623308	54926289	47.463	41.667
21)	B Endrin ke...	7.068	7.669	104.4E6	135.6E6	53.008	48.610
22)	Mirex	7.255	8.122	74388431	94826512	50.128	45.063

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101720\
Data File : PL062401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2020 17:33
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 02:37:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101320.M
Quant Title : GC Extractables
QLast Update : Tue Oct 13 05:04:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

