

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101823\
 Data File : PL085989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2023 11:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/19/2023
 Supervised By :Ankita Jodhani 10/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 13:13:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	2.622	132.7E6	39833163	44.629	42.643
28)	SA Decachlor...	8.860	7.774	118.7E6	51777933	54.500	56.098
Target Compounds							
2)	A alpha-BHC	3.813	3.121	193.8E6	55650853	45.161	44.233
3)	MA gamma-BHC...	4.145	3.448	190.5E6	56208551	46.081	45.574
4)	MA Heptachlor	4.732	3.784	188.0E6	55407130	48.602	47.639
5)	MB Aldrin	5.072	4.061	183.2E6	54294369	47.257	48.552
6)	B beta-BHC	4.347	3.750	81124343	24886425	49.059	47.804
7)	B delta-BHC	4.590	3.975	185.1E6	54995545	48.279	47.865
8)	B Heptachlo...	5.499	4.566	159.9E6	50434608	46.348m	50.214
9)	A Endosulfan I	5.883	4.933	148.6E6	47194566	45.953m	43.994
10)	B gamma-Chl...	5.757	4.818	162.1E6	51475598	47.364m	51.975
11)	B alpha-Chl...	5.836	4.881	165.0E6	52424897	48.289m	48.064
12)	B 4,4'-DDE	6.021	5.080	146.7E6	46094917	52.458	56.136
13)	MA Dieldrin	6.161	5.198	156.1E6	49198812	47.496	50.709m
14)	MA Endrin	6.388	5.473	132.1E6	46537730	46.341m	51.766
15)	B Endosulfa...	6.613	5.770	135.7E6	45256415	49.251	47.591
16)	A 4,4'-DDD	6.538	5.636	110.3E6	38253311	54.766m	54.958
17)	MA 4,4'-DDT	6.854	5.887	119.1E6	43494514	50.507	55.004
18)	B Endrin al...	6.744	5.952	104.8E6	38479816	51.725	51.602
19)	B Endosulfa...	6.979	6.176	126.5E6	45997762	50.405	50.989
20)	A Methoxychlor	7.340	6.472	67764273	26918193	48.520	51.168
21)	B Endrin ke...	7.461	6.678	133.0E6	55131027	46.165	47.964
22)	Mirex	7.934	6.855	107.2E6	47354791	48.403	50.208

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101823\
 Data File : PL085989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2023 11:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/19/2023

Supervised By :Ankita Jodhani 10/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 13:13:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

