

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101823\
 Data File : PL086016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2023 20:33
 Operator : AR\AJ
 Sample : 04938-11
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 A-6(0-2)

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/19/2023
 Supervised By :Ankita Jodhani 10/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 22:25:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.355	2.622	32810499	10237111	11.032	10.959
2) SA Decachlor...	8.868	7.771	9203773	12067859	4.226m	13.075 #
Target Compounds						
12) B 4,4'-DDE	6.016	5.080	4030213	717337	1.441m	0.874m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101823\
Data File : PL086016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2023 20:33
Operator : AR\AJ
Sample : 04938-11
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
A-6(0-2)


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2023
Supervised By :Ankita Jodhani 10/19/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 22:25:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
Quant Title : GC Extractables
QLast Update : Wed Oct 04 05:29:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.355 min
Delta R.T.: -0.004 min
Response: 32810499
Conc: 11.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
A-6(0-2)


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2023
Supervised By :Ankita Jodhani 10/19/2023




#1 Tetrachloro-m-xylene

R.T.: 2.622 min
Delta R.T.: 0.000 min
Response: 10237111
Conc: 10.96 ng/ml




#12 4,4'-DDE

R.T.: 6.016 min
Delta R.T.: -0.005 min
Response: 4030213
Conc: 1.44 ng/ml m



#12 4,4'-DDE

R.T.: 5.080 min
Delta R.T.: -0.002 min
Response: 717337
Conc: 0.87 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.868 min
Delta R.T.: 0.009 min
Response: 9203773
Conc: 4.23 ng/ml

Instrument :
ECD_L
ClientSampleId :
A-6(0-2)

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2023
Supervised By :Ankita Jodhani 10/19/2023



#28 Decachlorobiphenyl

R.T.: 7.771 min
Delta R.T.: -0.003 min
Response: 12067859
Conc: 13.07 ng/ml