

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(10) trans-Chlordane (B)

5.916min 0.528 ng/ml

response 1027732

(10) trans-Chlordane #2 (B)

5.013min 1.460 ng/ml

response 3517200

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(10) trans-Chlordane (B)

5.944min 0.534 ng/ml m

response 1040182

(10) trans-Chlordane #2 (B)

5.011min 0.442 ng/ml m

response 1065646

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(15) Endosulfan II (B)

6.823min 1.704 ng/ml

response 3020204

(15) Endosulfan II #2 (B)

5.928min 0.624 ng/ml

response 1321749

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(15) Endosulfan II (B)
6.821min 1.437 ng/ml m
response 2547455

(15) Endosulfan II #2 (B)
5.927min 0.508 ng/ml m
response 1076796

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.711min 1.037 ng/ml
response 1478484

(16) 4,4'-DDD #2 (A)

5.792min 0.443 ng/ml
response 810019

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.709min 1.102 ng/ml m

response 1571407

(16) 4,4'-DDD #2 (A)

5.791min 0.819 ng/ml m

response 1497679

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(27) Decachlorobiphenyl (SA)

9.054min 6.835 ng/ml

response 9184513

(27) Decachlorobiphenyl #2 (SA)

7.917min 6.999 ng/ml

response 14892365

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101824\
Data File : PL092468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:09
Operator : AR\AJ
Sample : P4346-13
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
EOAX3

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 07:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(27) Decachlorobiphenyl (SA)

9.052min 7.260 ng/ml m

response 9754782

(27) Decachlorobiphenyl #2 (SA)

7.917min 6.999 ng/ml

response 14892365