

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101919\
 Data File : PL053484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Oct 2019 00:44
 Operator : SG\AJ
 Sample : K5420-03MSD
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-2MSD

Manual Integrations
 APPROVED

Ankita
 10/21/2019 12:12:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 01:45:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	4.045	192.0E6	47466077	17.711	18.968
28)	SA Decachlor...	8.332	9.125	150.7E6	40651585	15.363	14.436
Target Compounds							
2)	A alpha-BHC	3.988	4.437	959.3E6	198.4E6	55.016	57.400
3)	MA gamma-BHC...	4.275	4.721	864.1E6	187.1E6	55.488	58.091m
4)	MA Heptachlor	4.582	5.232	854.9E6	174.7E6	54.167	54.203
5)	MB Aldrin	4.834	5.538	769.7E6	164.7E6	53.840	55.230
6)	B beta-BHC	4.524	4.891	347.4E6	80378222	52.712	52.037
7)	B delta-BHC	4.733	5.110	754.3E6	150.5E6	49.553	49.818
8)	B Heptachlor...	5.280	5.924	711.6E6	155.2E6	52.713	53.224
9)	A Endosulfan I	5.622	6.281	621.2E6	140.6E6	50.533	51.868
10)	B gamma-Chl...	5.508	6.158	771.0E6	157.7E6	56.510	54.061
11)	B alpha-Chl...	5.567	6.231	731.1E6	164.0E6	54.953	56.181
12)	B 4,4'-DDE	5.731	6.386	612.5E6	142.6E6	51.549	50.623
13)	MA Dieldrin	5.867	6.538	658.9E6	145.5E6	50.553	50.983
14)	MA Endrin	6.126	6.756	584.0E6	124.1E6	50.379	51.120
15)	B Endosulfa...	6.402	6.969	970.7E6	246.7E6	86.824	96.563
16)	A 4,4'-DDD	6.248	6.876	450.3E6	112.5E6	50.592	53.976
17)	MA 4,4'-DDT	6.491	7.175	509.6E6	119.8E6	51.442	51.790
18)	B Endrin al...	6.567	7.089	456.8E6	113.2E6	49.816	51.521
19)	B Endosulfa...	6.780	7.311	480.2E6	120.5E6	47.206	49.109
20)	A Methoxychlor	7.036	7.633	235.3E6	62019347	47.731	47.747
21)	B Endrin ke...	7.274	7.782	538.6E6	136.9E6	50.031	53.077
22)	Mirex	7.468	8.238	445.2E6	110.9E6	52.048	54.838

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101919\
Data File : PL053484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Oct 2019 00:44
Operator : SG\AJ
Sample : K5420-03MSD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
TP-2MSD

Manual Integrations
APPROVED

Ankita
10/21/2019 12:12:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 01:45:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

