

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102020\
 Data File : PL062446.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2020 15:05
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 04:48:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101920.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 20 07:00:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.411	3.950	88262300	118.7E6	51.227	51.538
2)	SA Decachlor...	8.125	8.987	82059376	95742322	46.749	48.284
Target Compounds							
2)	A alpha-BHC	3.843	4.336	128.9E6	173.5E6	51.545	51.529
3)	MA gamma-BHC...	4.122	4.619	115.9E6	162.3E6	51.589	51.348
4)	MA Heptachlor	4.419	5.125	107.7E6	152.7E6	51.582	51.105
5)	MB Aldrin	4.663	5.428	95860475	138.5E6	51.433	50.446
6)	B beta-BHC	4.370	4.790	48284835	68641673	51.766	47.538
7)	B delta-BHC	4.571	5.005	110.0E6	144.0E6	51.270	55.178
8)	B Heptachlo...	5.101	5.813	91314967	126.7E6	51.371	52.434
9)	A Endosulfan I	5.436	6.168	83901112	117.3E6	51.283	48.537
10)	B gamma-Chl...	5.326	6.046	90574690	126.2E6	51.428	49.826
11)	B alpha-Chl...	5.384	6.119	90142246	126.5E6	50.921	49.078
12)	B 4,4'-DDE	5.549	6.278	82585725	118.5E6	51.122	48.753
13)	MA Dieldrin	5.677	6.424	87861303	119.5E6	51.103	48.713
14)	MA Endrin	5.931	6.641	81113923	103.5E6	49.966	49.450
15)	B Endosulfa...	6.203	6.851	81245376	104.0E6	49.368	47.490
16)	A 4,4'-DDD	6.060	6.766	70574500	93088340	50.823	48.815
17)	MA 4,4'-DDT	6.299	7.065	68990334	90989386	51.354	48.009
18)	B Endrin al...	6.370	6.974	69306371	88090996	48.820	47.137
19)	B Endosulfa...	6.582	7.198	80586781	97569782	50.154	47.544
20)	A Methoxychlor	6.843	7.524	42749599	53129773	50.380	47.895
21)	B Endrin ke...	7.069	7.667	91753887	103.2E6	49.548	48.484
22)	Mirex	7.255	8.119	70217268	77122496	47.818	47.084

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102020\
Data File : PL062446.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2020 15:05
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 04:48:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101920.M
Quant Title : GC Extractables
QLast Update : Tue Oct 20 07:00:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

