

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102022\
 Data File : PL078426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2022 11:30
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 12:34:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101222.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 12 15:30:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.418	2.619	76036215	129.6E6	46.009	49.036
2)	SA Decachlor...	8.909	7.717	64615563	104.0E6	41.855	47.842
Target Compounds							
2)	A alpha-BHC	3.884	3.112	118.6E6	204.4E6	47.147	49.116
3)	MA gamma-BHC...	4.220	3.436	107.2E6	184.7E6	45.981	53.002
4)	MA Heptachlor	4.794	3.767	101.3E6	181.0E6	45.562	50.679
5)	MB Aldrin	5.135	4.041	108.4E6	170.5E6	51.765	50.114
6)	B beta-BHC	4.433	3.736	45501432	74202731	45.748	50.332
7)	B delta-BHC	4.677	3.959	106.3E6	182.8E6	57.829	49.441
8)	B Heptachlo...	5.569	4.540	99825097	153.3E6	65.567	52.023
9)	A Endosulfan I	5.954	4.906	84459155	142.8E6	50.311	51.792
10)	B gamma-Chl...	5.824	4.789	96778259	154.1E6	54.315	51.484
11)	B alpha-Chl...	5.901	4.852	91437873	149.1E6	51.061	51.792
12)	B 4,4'-DDE	6.081	5.046	96037481	148.8E6	56.983	52.063
13)	MA Dieldrin	6.232	5.169	85116942	151.2E6	49.472	51.595
14)	MA Endrin	6.462	5.441	68917511	132.2E6	46.185	57.396
15)	B Endosulfa...	6.692	5.738	67778774	121.7E6	50.559	49.717
16)	A 4,4'-DDD	6.606	5.599	64681349	122.8E6	52.530	55.139
17)	MA 4,4'-DDT	6.915	5.847	69441956	118.0E6	47.435	51.973
18)	B Endrin al...	6.823	5.917	51538889	93214285	45.036	50.377
19)	B Endosulfa...	7.056	6.141	63912820	120.4E6	44.809	50.032
20)	A Methoxychlor	7.397	6.425	30983285	61188080	40.891	48.446
21)	B Endrin ke...	7.544	6.639	73172459	136.0E6	46.903	51.341
22)	Mirex	8.003	6.815	53111691	95480095	44.966	49.630

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102022\
Data File : PL078426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2022 11:30
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 12:34:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101222.M
Quant Title : GC Extractables
QLast Update : Wed Oct 12 15:30:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

