

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
 Data File : PL053510.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 12:05
 Operator : SG\AJ
 Sample : K5111-02
 Misc : CHLOR LOQ 50 PPB
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 LOQ-SOIL-02-QT4-2019

Manual Integrations
 APPROVED

Ankita
 10/22/2019 12:23:50 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 02:14:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	4.045	186.6E6	44021220	17.217	17.592
28)	SA Decachlor...	8.330	9.123	176.3E6	51577397	17.971	18.316
Target Compounds							
23)	Chlordane-1	4.427	5.042	20236098	4545683	57.056	54.201
24)	Chlordane-2	4.933	5.510	22435019	5249086	58.225	53.441
25)	Chlordane-3	5.507	6.155	52426498	16178432	45.482	44.297m
26)	Chlordane-4	5.566	6.230	45401192	20556786	48.430	45.790
27)	Chlordane-5	6.406	7.036	14967332	4606891	40.945	60.262 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
Data File : PL053510.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 12:05
Operator : SG\AJ
Sample : K5111-02
Misc : CHLOR LOQ 50 PPB
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
LOQ-SOIL-02-QT4-2019

Manual Integrations
APPROVED

Ankita
10/22/2019 12:23:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 02:14:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

