

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
 Data File : PL053527.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 16:03
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 01:01:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	4.045	510.9E6	119.6E6	47.136	47.805
2)	SA Decachlor...	8.331	9.123	473.3E6	133.8E6	48.239	47.497
Target Compounds							
2)	A alpha-BHC	3.987	4.436	838.8E6	170.2E6	48.105	49.222
3)	MA gamma-BHC...	4.274	4.721	756.6E6	158.1E6	48.585	49.093
4)	MA Heptachlor	4.581	5.231	752.0E6	153.1E6	47.644	47.510
5)	MB Aldrin	4.833	5.536	701.1E6	144.6E6	49.040	48.507
6)	B beta-BHC	4.524	4.891	307.1E6	72288635	46.603	46.800
7)	B delta-BHC	4.732	5.108	732.4E6	150.6E6	48.116	49.848
8)	B Heptachlo...	5.278	5.922	646.4E6	139.3E6	47.882	47.753
9)	A Endosulfan I	5.621	6.279	592.4E6	131.2E6	48.197	48.390
10)	B gamma-Chl...	5.507	6.156	655.3E6	141.7E6	48.033	48.548
11)	B alpha-Chl...	5.566	6.229	636.2E6	141.3E6	47.820	48.409
12)	B 4,4'-DDE	5.729	6.385	577.1E6	137.0E6	48.571	48.631
13)	MA Dieldrin	5.865	6.536	634.8E6	139.2E6	48.708	48.775
14)	MA Endrin	6.124	6.754	541.1E6	113.8E6	46.682	46.883
15)	B Endosulfa...	6.397	6.963	535.4E6	126.6E6	47.885	49.577
16)	A 4,4'-DDD	6.247	6.875	449.3E6	107.8E6	50.480	51.744
17)	MA 4,4'-DDT	6.489	7.174	468.7E6	109.5E6	47.310	47.369
18)	B Endrin al...	6.565	7.087	431.3E6	106.5E6	47.036	48.482
19)	B Endosulfa...	6.779	7.310	490.1E6	118.4E6	48.182	48.244
20)	A Methoxychlor	7.035	7.631	232.2E6	60279509	47.098	46.408
21)	B Endrin ke...	7.273	7.781	548.6E6	128.5E6	50.962	49.819
22)	Mirex	7.467	8.236	403.6E6	99668064	47.187	49.266

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
Data File : PL053527.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 16:03
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:01:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

