

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
 Data File : PL053553.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 22:08
 Operator : SG\AJ
 Sample : K5448-01MS
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SPMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 01:06:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	4.045	214.3E6	53797729	19.771	21.499
2)	SA Decachlor...	8.330	9.121	148.6E6	46390922	15.147	16.474
Target Compounds							
2)	A alpha-BHC	3.986	4.436	898.5E6	186.0E6	51.529	53.813
3)	MA gamma-BHC...	4.273	4.721	773.0E6	171.6E6	49.633	53.283
4)	MA Heptachlor	4.580	5.231	762.0E6	166.2E6	48.277	51.574
5)	MB Aldrin	4.832	5.536	707.1E6	151.5E6	49.464	50.826
6)	B beta-BHC	4.523	4.891	313.3E6	79373895	47.546	51.387
7)	B delta-BHC	4.731	5.108	672.9E6	139.2E6	44.210	46.081
8)	B Heptachlo...	5.278	5.922	647.1E6	143.4E6	47.933	49.164
9)	A Endosulfan I	5.620	6.279	577.0E6	127.8E6	46.941	47.140
10)	B gamma-Chl...	5.506	6.156	645.0E6	139.2E6	47.273	47.694
11)	B alpha-Chl...	5.565	6.229	620.9E6	137.2E6	46.670	47.003
12)	B 4,4'-DDE	5.728	6.384	584.4E6	140.4E6	49.179	49.846
13)	MA Dieldrin	5.864	6.536	594.7E6	137.8E6	45.628	48.294
14)	MA Endrin	6.123	6.753	515.1E6	113.2E6	44.441	46.645
15)	B Endosulfa...	6.396	6.962	479.5E6	124.0E6	42.888	48.541
16)	A 4,4'-DDD	6.246	6.874	412.5E6	106.6E6	46.353	51.159
17)	MA 4,4'-DDT	6.488	7.173	432.3E6	111.3E6	43.636	48.136
18)	B Endrin al...	6.564	7.086	373.1E6	99362804	40.692	45.231
19)	B Endosulfa...	6.778	7.309	421.4E6	108.6E6	41.426	44.259
20)	A Methoxychlor	7.033	7.630	208.1E6	59917158	42.218	46.129
21)	B Endrin ke...	7.272	7.780	466.6E6	121.7E6	43.347	47.210
22)	Mirex	7.465	8.235	398.6E6	103.6E6	46.600	51.195

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102119\
Data File : PL053553.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 22:08
Operator : SG\AJ
Sample : K5448-01MS
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
SPMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:06:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

