

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102122\
 Data File : PL078435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2022 22:25
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 06:37:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 06:37:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.323	2.573	138.9E6	181.1E6	73.559	73.205
28)	SA Decachlor...	8.777	7.644	106.0E6	128.5E6	72.503	72.450
Target Compounds							
2)	A alpha-BHC	3.784	3.060	219.9E6	289.2E6	73.382	73.512
3)	MA gamma-BHC...	4.116	3.381	202.2E6	258.0E6	73.439	73.313
4)	MA Heptachlor	4.687	3.708	192.8E6	248.2E6	73.450	73.277
5)	MB Aldrin	5.026	3.980	185.5E6	247.8E6	73.337	73.300
6)	B beta-BHC	4.331	3.681	82256516	99541350	72.834	73.441
7)	B delta-BHC	4.571	3.901	177.0E6	261.8E6	73.858	73.198
8)	B Heptachlo...	5.459	4.477	153.9E6	211.1E6	73.052	73.175
9)	A Endosulfan I	5.843	4.840	153.7E6	201.4E6	72.974	73.116
10)	B gamma-Chl...	5.715	4.725	165.5E6	215.0E6	72.945	73.151
11)	B alpha-Chl...	5.791	4.787	164.8E6	208.6E6	72.932	73.075
12)	B 4,4'-DDE	5.975	4.982	147.8E6	195.3E6	72.850	73.085
13)	MA Dieldrin	6.121	5.102	159.8E6	210.5E6	73.067	73.005
14)	MA Endrin	6.352	5.373	136.5E6	185.2E6	72.600	73.024
15)	B Endosulfa...	6.581	5.669	134.4E6	164.1E6	72.618	73.026
16)	A 4,4'-DDD	6.499	5.533	121.2E6	157.4E6	72.787	72.904
17)	MA 4,4'-DDT	6.809	5.780	131.4E6	161.7E6	72.846	73.160
18)	B Endrin al...	6.713	5.848	104.0E6	129.4E6	72.698	72.672
19)	B Endosulfa...	6.945	6.070	131.3E6	165.2E6	72.747	72.789
20)	A Methoxychlor	7.294	6.357	64972622	80702742	72.762	72.788
21)	B Endrin ke...	7.432	6.566	138.4E6	183.1E6	72.753	72.820
22)	Mirex	7.890	6.740	103.4E6	133.6E6	72.694	72.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102122\
Data File : PL078435.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2022 22:25
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 06:37:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 06:37:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

