

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102124\
 Data File : PL092514.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2024 18:08
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL102124

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:18:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX102124.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 21 17:30:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.778	123.7E6	134.1E6	47.255	48.670
7) SA Decachlor...	9.056	7.918	96187470	128.7E6	48.550	48.370
Target Compounds						
2) Toxaphene-1	6.238	5.007	12867496	12289931	482.998	487.266
3) Toxaphene-2	6.442	5.333	8135668	10574213	519.010	439.000
4) Toxaphene-3	7.061	6.606	41627340	32898788	509.733	478.723
5) Toxaphene-4	7.151	6.734	33399134	44775434	480.771	537.329
6) Toxaphene-5	7.936	7.047	23214509	33265320	465.716	426.029

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102124\
Data File : PL092514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 18:08
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL102124

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:18:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX102124.M
Quant Title : GC Extractables
QLast Update : Mon Oct 21 17:30:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x 0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

