

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102219\
 Data File : PL053586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2019 15:50
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/23/2019 11:13:01 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 01:04:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	4.045	566.0E6	129.3E6	52.221	51.665
28)	SA Decachlor...	8.334	9.126	441.3E6	132.4E6	44.974	47.000
Target Compounds							
2)	A alpha-BHC	3.990	4.436	894.7E6	184.3E6	51.312	53.306
3)	MA gamma-BHC...	4.277	4.722	778.6E6	170.9E6	49.994	53.055
4)	MA Heptachlor	4.585	5.232	773.1E6	161.8E6	48.983	50.215
5)	MB Aldrin	4.836	5.538	724.0E6	151.8E6	50.647	50.929
6)	B beta-BHC	4.528	4.890	314.3E6	78650861	47.701	50.919m
7)	B delta-BHC	4.736	5.110	763.6E6	161.3E6	50.165	53.384
8)	B Heptachlor...	5.282	5.924	664.7E6	141.0E6	49.241	48.338
9)	A Endosulfan I	5.624	6.281	587.0E6	129.1E6	47.751	47.601
10)	B gamma-Chl...	5.511	6.157	649.1E6	141.9E6	47.577	48.643
11)	B alpha-Chl...	5.570	6.231	625.6E6	140.4E6	47.028	48.090
12)	B 4,4'-DDE	5.733	6.387	587.1E6	137.8E6	49.413	48.931
13)	MA Dieldrin	5.869	6.538	621.3E6	139.7E6	47.669	48.943
14)	MA Endrin	6.128	6.756	541.1E6	114.9E6	46.683	47.349
15)	B Endosulfa...	6.400	6.965	500.6E6	144.2E6	44.779	56.441 #
16)	A 4,4'-DDD	6.250	6.877	472.1E6	114.1E6	53.043	54.734
17)	MA 4,4'-DDT	6.493	7.175	433.5E6	104.9E6	43.764	45.345
18)	B Endrin al...	6.569	7.089	411.9E6	102.3E6	44.922	46.555
19)	B Endosulfa...	6.782	7.311	466.8E6	115.9E6	45.891	47.234
20)	A Methoxychlor	7.038	7.633	216.9E6	57452678	43.996	44.231
21)	B Endrin ke...	7.276	7.782	515.1E6	126.9E6	47.851	49.215
22)	Mirex	7.470	8.238	366.9E6	92592782	42.901	45.769

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102219\
Data File : PL053586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2019 15:50
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/23/2019 11:13:01 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 01:04:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

