

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102220\
 Data File : PL062537.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2020 20:07
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 04:52:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.413	3.952	92814458	162.8E6	53.959	55.392
28)	SA Decachlor...	8.127	8.990	79354809	110.8E6	49.464	52.958
Target Compounds							
2)	A alpha-BHC	3.844	4.338	139.1E6	252.5E6	54.571	56.821
3)	MA gamma-BHC...	4.124	4.622	124.3E6	233.0E6	53.635	55.806
4)	MA Heptachlor	4.420	5.126	113.1E6	213.6E6	52.713	53.555
5)	MB Aldrin	4.665	5.429	103.3E6	200.5E6	54.143	55.399
6)	B beta-BHC	4.373	4.792	51563724	100.1E6	53.983	53.156
7)	B delta-BHC	4.573	5.007	118.7E6	204.1E6	54.631	58.312
8)	B Heptachlo...	5.103	5.815	100.3E6	178.7E6	54.946	55.173
9)	A Endosulfan I	5.438	6.170	89381836	164.3E6	53.470	52.106
10)	B gamma-Chl...	5.329	6.048	95420697	179.2E6	52.965	53.241
11)	B alpha-Chl...	5.386	6.121	95333066	177.4E6	52.735	52.661
12)	B 4,4'-DDE	5.551	6.279	89019190	168.3E6	53.806	52.250
13)	MA Dieldrin	5.679	6.426	94460896	166.2E6	53.459	51.912
14)	MA Endrin	5.934	6.643	82809182	134.7E6	49.772	48.252
15)	B Endosulfa...	6.206	6.853	85623871	140.0E6	51.936	48.959
16)	A 4,4'-DDD	6.063	6.769	80622274	137.4E6	56.997	53.668
17)	MA 4,4'-DDT	6.302	7.067	61512623	106.4E6	46.092	47.383
18)	B Endrin al...	6.374	6.977	74189163	119.1E6	51.092	52.204
19)	B Endosulfa...	6.586	7.200	83398492	128.6E6	50.537	53.612
20)	A Methoxychlor	6.846	7.526	38528774	59544425	44.734	48.194
21)	B Endrin ke...	7.072	7.668	95242292	134.3E6	52.824	56.662
22)	Mirex	7.258	8.121	69020260	90441981	48.685	53.074

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102220\
Data File : PL062537.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2020 20:07
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 04:52:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

